

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3787

Analytical Method: SW8260-Low

Client: JACOBS Engineering Group, Inc

Datafile : VX048764.D

											Limits		
Parameter	Spike	Sample	Result	Units	Rec			RPD	Qual	Low	High	RPD	
		Result			Rec	Qual	RPD						
Lab Sample ID :	Q3787-02MS	Client Sample ID :	MW-15B-42.5-120425-MS										
Vinyl chloride	50	1.30	51.8	ug/L	101					70 (69)	130 (125)		
1,1-Dichloroethene	50	0	49.2	ug/L	98					70 (80)	130 (124)		
1,1-Dichloroethane	50	0	58.8	ug/L	118					70 (78)	130 (122)		
cis-1,2-Dichloroethene	50	3.30	59.1	ug/L	112					70 (55)	130 (150)		
1,1,1-Trichloroethane	50	0	51.6	ug/L	103					70 (83)	130 (117)		
Benzene	50	0	49.9	ug/L	100					70 (81)	130 (128)		
1,2-Dichloroethane	50	0	52.7	ug/L	105					70 (76)	130 (120)		
Trichloroethene	50	6.20	52.5	ug/L	93					70 (44)	130 (168)		
1,1,2-Trichloroethane	50	0	51.4	ug/L	103					70 (73)	130 (136)		
Tetrachloroethene	50	0	38.7	ug/L	77					70 (63)	130 (122)		