

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3787

Analytical Method: SW8260-Low

Client: JACOBS Engineering Group, Inc.

Datafile : VX048765.D

Parameter	Spike	Sample	Result	Units	Rec		RPD	Limits			
		Result			Qual	RPD	Qual	Low	High	RPD	
Lab Sample ID :	Q3787-03MSD	Client Sample ID :	MW-15B-42.5-120425-MSD								
Vinyl chloride	50	1.30	50.9	ug/L	99		2		70 (69)	130 (125)	20 (20)
1,1-Dichloroethene	50	0	49.7	ug/L	99		1		70 (80)	130 (124)	20 (20)
1,1-Dichloroethane	50	0	60.4	ug/L	121		3		70 (78)	130 (122)	20 (20)
cis-1,2-Dichloroethene	50	3.30	60.7	ug/L	115		3		70 (55)	130 (150)	20 (20)
1,1,1-Trichloroethane	50	0	53.2	ug/L	106		3		70 (83)	130 (117)	20 (20)
Benzene	50	0	50.1	ug/L	100		0		70 (81)	130 (128)	20 (20)
1,2-Dichloroethane	50	0	55.2	ug/L	110		5		70 (76)	130 (120)	20 (20)
Trichloroethene	50	6.20	52.7	ug/L	93		0		70 (44)	130 (168)	20 (20)
1,1,2-Trichloroethane	50	0	51.4	ug/L	103		0		70 (73)	130 (136)	20 (20)
Tetrachloroethene	50	0	39.5	ug/L	79		2		70 (63)	130 (122)	20 (20)