

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JACO05
Lab Code: ACE SDG NO.: Q3787
Lab File ID: VX048796.D Date Analyzed: 12/10/2025
Instrument ID: MSVOA_X Time Analyzed: 09:40
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	164636	5.53	275167	6.74	243543	10.04
UPPER LIMIT	329272	6.031	550334	7.238	487086	10.537
LOWER LIMIT	82318	5.031	137584	6.238	121772	9.537
EPA SAMPLE NO.						
OWBR-02-170-120425RE	160525	5.54	337243	6.75	371573	10.04
OWBR-02-170-120425-FDRE	150728	5.54	313869	6.75	341495	10.04
OW-03B-51.5-120425RE	142606	5.54	299203	6.75	326420	10.04
OW-03B-51.5-120425-FDRE	152091	5.54	320261	6.75	354570	10.04
OW-08B-72.5-120425-FDRE	154804	5.54	318388	6.75	343017	10.04
OW-02B-21.2-120425	138008	5.54	284336	6.75	306079	10.04
TB01-120425	137750	5.54	283821	6.75	308054	10.04
VX1210WBL01	143182	5.53	288257	6.74	306883	10.04
VX1210WBS01	153473	5.53	261026	6.74	236663	10.04

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.