

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>JAC005</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3787</u>
Lab File ID:	<u>VX048742.D</u>	Date Analyzed:	<u>12/08/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>09:20</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	132353	12				
UPPER LIMIT	264706	12.5				
LOWER LIMIT	66176.5	11.5				
EPA SAMPLE NO.						
MW-15B-42.5-120425	129079	12.01				
MW-15B-42.5-120425-MS	125815	12.01				
MW-15B-42.5-120425-MSD	103730	12.01				
OWBR-02-170-120425	191518	12.01				
OWBR-02-170-120425-FD	176527	12.01				
OW-03B-51.5-120425	128068	12.01				
OW-03B-51.5-120425-FD	118375	12.01				
OW-08B-72.5-120425	119855	12.01				
OW-08B-72.5-120425-FD	171945	12.01				
VX1208WBL01	179749	12.01				
VX1208WBS01	157846	12.00				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.