

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JACO05
 Lab Code: ACE SDG NO.: Q3787
 Lab File ID: VX048796.D Date Analyzed: 12/10/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:40
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	117430	12				
UPPER LIMIT	234860	12.5				
LOWER LIMIT	58715	11.5				
EPA SAMPLE NO.						
OWBR-02-170-120425RE	199241	12.01				
OWBR-02-170-120425-FDRE	184829	12.01				
OW-03B-51.5-120425RE	172615	12.01				
OW-03B-51.5-120425-FDRE	184719	12.01				
OW-08B-72.5-120425-FDRE	186375	12.01				
OW-02B-21.2-120425	165546	12.01				
TB01-120425	167960	12.01				
VX1210WBL01	162084	12.00				
VX1210WBS01	123629	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.