

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>JAC005</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3788</u>
Instrument ID:	<u>BNA_N</u>	Calibration Date/Time:	<u>12/12/2025 09:24</u>
Lab File ID:	<u>BN038343.D</u>	Init. Calib. Date(s):	<u>12/10/2025 12/10/2025</u>
EPA Sample No.:	<u>SSTDCCC0.4</u>	Init. Calib. Time(s):	<u>09:52 13:30</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.564	0.557		-1.2	20.0
Fluoranthene-d10	0.990	1.007		1.7	20.0
2-Fluorophenol	0.996	0.966		-3.0	20.0
Phenol-d6	1.184	1.162		-1.9	20.0
Nitrobenzene-d5	0.337	0.317		-5.9	20.0
2-Fluorobiphenyl	1.914	1.930		0.8	20.0
2,4,6-Tribromophenol	0.177	0.159		-10.2	20.0
Terphenyl-d14	0.893	0.853		-4.5	20.0
1,4-Dioxane	0.459	0.488		6.3	20.0

All other compounds must meet a minimum RRF of 0.010.