

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3788

Analytical Method: SW8270-Modified

Client: JACOBS Engineering Group, Inc.

DataFile: BN038336.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q3788-03MSD		Client Sample ID: OWBR-05-135-120525MSD									
1,4-Dioxane	0.4	34.3	28.3	ug/L	-1500	*	60	*	20 (10)	160 (175)	20 (20)