

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3788

Client ID: SSTDCCC0.4

Date Analyzed: 12/12/2025

Lab File ID: BN038343.D

Time Analyzed: 09:24

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	14177	17.148	11670	21.339	12444	23.633
UPPER LIMIT	28354	17.648	23340	21.839	24888	24.133
LOWER LIMIT	7088.5	16.648	5835	20.839	6222	23.133
EPA SAMPLE NO.						
01 OWBR-05-135-120525DL	16300	17.15	12712	21.34	12609	23.63

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.