

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN038329.D
Instrument ID: BNA_N

Contract: JAC005
SDG NO.: Q3788
DFTPP Injection Date: 12/11/2025
DFTPP Injection Time: 23:20

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	37.8
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.1
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	15.3
442	Greater than 50% of mass 198	95.1
443	15.0 - 24.0% of mass 442	18.9 (19.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN038330.D	12/11/2025	23:59
OWBR-05-135-120525	Q3788-01	BN038334.D	12/12/2025	02:25
OWBR-05-135-120525MS	Q3788-02MS	BN038335.D	12/12/2025	03:01
OWBR-05-135-120525MSD	Q3788-03MSD	BN038336.D	12/12/2025	03:37
BR-05-465-120525	Q3788-09	BN038340.D	12/12/2025	06:02