

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120825\
 Data File : VV039496.D
 Acq On : 08 Dec 2025 16:24
 Operator : SY/MD
 Sample : Q3788-05MS
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 OWBR-05-135-120525-SIMMS

Quant Time: Dec 09 01:28:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Dec 04 00:58:46 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	7409	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.784	117	7334	0.500	ug/L	0.00
12) 1,4-Dichlorobenzene-d4	11.175	152	4127	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.294	65	3975	0.534	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 106.000%	
4) 1,2-Dichloroethane-d4	5.017	65	3997	0.810	ug/L	0.05
Spiked Amount	0.500	Range	70 - 130	Recovery	= 162.000%#	
7) 1,2-Dichloropropane-d6	6.019	67	3475	0.562	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 112.000%	
8) Toluene-d8	7.253	98	5556	0.532	ug/L	0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	= 106.000%	
10) 1,1,2,2-Tetrachloroeth...	10.132	84	1891	0.542	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 108.000%	
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.294	62	17638	1.100	ug/L #	1
6) Trichloroethene	5.828	95	200661	22.075	ug/L	96
9) 1,2-Dibromoethane	8.289	107	2875	0.479	ug/L	99
11) 1,1,2,2-Tetrachloroethane	10.154	83	3888	0.528	ug/L	100
13) 1,2,3-Trichloropropane	10.199	75	2192	0.505	ug/L	97
14) 1,2-Dibromo-3-chloropr...	12.350	75	426	0.484	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120825\
Data File : VV039496.D
Acq On : 08 Dec 2025 16:24
Operator : SY/MD
Sample : Q3788-05MS
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
OWBR-05-135-120525-SIMMS

Quant Time: Dec 09 01:28:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Dec 04 00:58:46 2025
Response via : Initial Calibration

