

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JACO05
Lab Code: ACE SDG NO.: Q3788
Lab File ID: VV039486.D Date Analyzed: 12/08/2025
Instrument ID: MSVOA_V Time Analyzed: 09:00
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 (DFB) AREA #	RT #	IS2 (CBZ) AREA #	RT #	IS3 (DCB) AREA #	RT #
12 HOUR STD	6968	5.57	7021	8.79	3931	11.18
UPPER LIMIT	13936	5.736	14042	8.955	7862	11.345
LOWER LIMIT	3484	5.396	3510.5	8.615	1965.5	11.005
EPA SAMPLE NO.						
OWBR-05-135-120525-SIM	6770	5.57	6255	8.80	3155	11.20
OWBR-05-135-120525-SIMMS	7409	5.57	7334	8.78	4127	11.18
OWBR-05-135-120525-SIMMSD	7167	5.57	7036	8.79	3963	11.18
TB01-120525	5977	5.57	5887	8.80	3057	11.20
VBLK237	7024	5.57	6420	8.79	3401	11.20

IS1 (DFB) = 1,4-Difluorobenzene

IS2 (CBZ) = Chlorobenzene-d5

IS3 (DCB) = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.