



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121025\
Data File : VX048795.D
Acq On : 10 Dec 2025 09:13
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Title : SW846 8260
Last Update : Fri Dec 05 03:27:34 2025



AutoFind: Scans 1635, 1636, 1637; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.0	6664	PASS
75	95	30	60	50.9	17843	PASS
95	95	100	100	100.0	35035	PASS
96	95	5	9	7.4	2579	PASS
173	174	0.00	2	0.7	200	PASS
174	95	50	100	77.6	27176	PASS
175	174	5	9	7.4	2018	PASS
176	174	95	101	97.1	26387	PASS
177	176	5	9	6.7	1759	PASS