

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX120425

Review By	John Carlone	Review On	12/5/2025 8:05:59 AM
Supervise By	Semsettin Yesilyurt	Supervise On	12/5/2025 2:14:46 PM
SubDirectory	VX120425	HP Acquire Method	HP Processing Method 82X120425W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP136497		
Initial Calibration Stds	VP136526,VP136527,VP136528,VP136529,VP136530,VP136531		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP136532		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048713.D	04 Dec 2025 08:15		JC/MD	Ok
2	VSTDICC001	VSTDICC001	VX048714.D	04 Dec 2025 08:47		JC/MD	Ok,M
3	VSTDICC005	VSTDICC005	VX048715.D	04 Dec 2025 09:16		JC/MD	Ok,M
4	VSTDICC020	VSTDICC020	VX048716.D	04 Dec 2025 09:36	Comp. #13 on Linear Regression	JC/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VX048717.D	04 Dec 2025 09:57	Comp. #95 on Quadratic Regression	JC/MD	Ok,M
6	VSTDICC100	VSTDICC100	VX048718.D	04 Dec 2025 10:17		JC/MD	Ok,M
7	VSTDICC150	VSTDICC150	VX048719.D	04 Dec 2025 10:38		JC/MD	Ok,M
8	IBLK	IBLK	VX048720.D	04 Dec 2025 10:58		JC/MD	Ok
9	VSTDICV050	ICVVX120425	VX048721.D	04 Dec 2025 14:04		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX120825

Review By	John Carlone	Review On	12/9/2025 8:11:51 AM
Supervise By	Maresh Dadoda	Supervise On	12/9/2025 12:45:16 PM
SubDirectory	VX120825	HP Acquire Method	HP Processing Method 82X120425W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136561
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136570,VP136571

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048741.D	08 Dec 2025 08:17		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX048742.D	08 Dec 2025 09:20	pH#Lot#V12668	JC/MD	Ok,M
3	VX1208MBL01	VX1208MBL01	VX048743.D	08 Dec 2025 09:49		JC/MD	Ok
4	VX1208WBL01	VX1208WBL01	VX048744.D	08 Dec 2025 11:15		JC/MD	Ok
5	VX1208WBS01	VX1208WBS01	VX048745.D	08 Dec 2025 11:44		JC/MD	Ok,M
6	VX1208WBSD01	VX1208WBSD01	VX048746.D	08 Dec 2025 12:12		JC/MD	Ok,M
7	Q3757-05RE	MW-14B-46-120325RE	VX048747.D	08 Dec 2025 12:33	vial B pH<2 Surrogate fail	JC/MD	Confirms
8	Q3757-09RE	OW-05B-72-120325RE	VX048748.D	08 Dec 2025 12:53	vial B pH<2 Surrogate fail	JC/MD	Confirms
9	Q3787-01	MW-15B-42.5-120425	VX048749.D	08 Dec 2025 13:14	vial A pH<2	JC/MD	Ok
10	Q3787-02MS	MW-15B-42.5-120425M	VX048750.D	08 Dec 2025 13:35	Not Spiked	JC/MD	Not Ok
11	Q3787-03MSD	MW-15B-42.5-120425M	VX048751.D	08 Dec 2025 13:55	Not Spiked	JC/MD	Not Ok
12	Q3787-07	OWBR-02-170-120425	VX048752.D	08 Dec 2025 14:16	vial A pH<2 Surrogate fail	JC/MD	ReRun
13	Q3787-09	OWBR-02-170-120425	VX048753.D	08 Dec 2025 14:36	vial A pH<2 Surrogate fail	JC/MD	ReRun
14	Q3787-11	OW-03B-51.5-120425	VX048754.D	08 Dec 2025 14:57	vial A pH<2 Surrogate fail	JC/MD	ReRun
15	Q3787-13	OW-03B-51.5-120425	VX048755.D	08 Dec 2025 15:18	vial A pH<2 Surrogate fail	JC/MD	ReRun
16	Q3787-15	OW-08B-72.5-120425	VX048756.D	08 Dec 2025 15:38	vial A pH<2	JC/MD	Ok
17	Q3787-17	OW-08B-72.5-120425	VX048757.D	08 Dec 2025 15:59	vial A pH<2 Surrogate fail	JC/MD	ReRun
18	Q3787-19	OW-02B-21.2-120425	VX048758.D	08 Dec 2025 16:19	vial A pH<2 Surrogate fail	JC/MD	ReRun

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX120825

Review By	John Carlone	Review On	12/9/2025 8:11:51 AM
Supervise By	Maresh Dadoda	Supervise On	12/9/2025 12:45:16 PM
SubDirectory	VX120825	HP Acquire Method	HP Processing Method 82X120425W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP136561		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136570,VP136571		

19	Q3787-21	TB01-120425	VX048759.D	08 Dec 2025 16:40	vial A pH<2 TB;Surrogate fail	JC/MD	ReRun
20	Q3788-07	OW-01B-66.5-120525	VX048760.D	08 Dec 2025 17:00	vial A pH<2 Surrogate fail	JC/MD	ReRun
21	Q3788-09	BR-05-465-120525	VX048761.D	08 Dec 2025 17:21	vial A pH<2 Surrogate fail	JC/MD	ReRun
22	Q3788-11	MW-16B-87.5-120525	VX048762.D	08 Dec 2025 17:41	vial A pH<2 Surrogate fail	JC/MD	ReRun
23	Q3788-13	TB01-120525	VX048763.D	08 Dec 2025 18:02	vial A pH<2 TB	JC/MD	Ok
24	Q3787-02MS	MW-15B-42.5-120425M	VX048764.D	08 Dec 2025 18:22	vial A pH<2	JC/MD	Ok,M
25	Q3787-03MSD	MW-15B-42.5-120425M	VX048765.D	08 Dec 2025 18:43	vial A pH<2	JC/MD	Ok,M
26	VSTDCCC050	VSTDCCC050EC	VX048766.D	08 Dec 2025 19:03		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX120925

Review By	John Carlone	Review On	12/10/2025 8:35:12 AM
Supervise By	Maresh Dadoda	Supervise On	12/10/2025 11:33:33 AM
SubDirectory	VX120925	HP Acquire Method	HP Processing Method 82X120425W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136583
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136584,VP136585

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048767.D	09 Dec 2025 08:16		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX048768.D	09 Dec 2025 09:25	pH#Lot#V12668	JC/MD	Ok,M
3	VX1209WBL01	VX1209WBL01	VX048769.D	09 Dec 2025 09:46		JC/MD	Ok
4	VX1209MBL01	VX1209MBL01	VX048770.D	09 Dec 2025 10:06		JC/MD	Ok
5	Q3803-01	CHASE-A	VX048771.D	09 Dec 2025 10:36	Need straight run	JC/MD	Not Ok
6	VX1209WBS01	VX1209WBS01	VX048772.D	09 Dec 2025 10:56	Recovery fail	JC/MD	Not Ok
7	VX1209WBS02	VX1209WBS02	VX048773.D	09 Dec 2025 11:24		JC/MD	Ok,M
8	Q3788-01	OWBR-05-135-120525	VX048774.D	09 Dec 2025 11:45	vial A pH<2	JC/MD	Ok
9	Q3788-02MS	OWBR-05-135-120525	VX048775.D	09 Dec 2025 12:05	Recovery Fail For Many Compund	JC/MD	Not Ok
10	Q3788-03MSD	OWBR-05-135-120525	VX048776.D	09 Dec 2025 12:25	vial A pH<2	JC/MD	Ok,M
11	Q3803-01	CHASE-A	VX048777.D	09 Dec 2025 12:46	vial A pH#5.0	JC/MD	Ok,M
12	Q3803-02	CHASE-B	VX048778.D	09 Dec 2025 13:06	vial A pH#5.0	JC/MD	Ok,M
13	Q3803-03	CHASE-C	VX048779.D	09 Dec 2025 13:26	vial A pH#5.0	JC/MD	Ok,M
14	Q3803-04	CHASE-D	VX048780.D	09 Dec 2025 13:47	vial A pH#5.0	JC/MD	Ok
15	Q3803-05	CHASE-E	VX048781.D	09 Dec 2025 14:07	vial A pH#5.0	JC/MD	Ok
16	Q3803-06	CHASE-F	VX048782.D	09 Dec 2025 14:27	vial A pH#5.0	JC/MD	Ok
17	Q3803-07	CHASE-G	VX048783.D	09 Dec 2025 14:48	vial A pH#5.0	JC/MD	Ok,M

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX120925

Review By	John Carlone	Review On	12/10/2025 8:35:12 AM
Supervise By	Maresh Dadoda	Supervise On	12/10/2025 11:33:33 AM
SubDirectory	VX120925	HP Acquire Method	HP Processing Method 82X120425W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP136583		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136584,VP136585		

18	Q3803-08	CHASE-H	VX048784.D	09 Dec 2025 15:08	vial A pH#5.0 Need 5X	JC/MD	Dilution
19	Q3803-09	CHASE-I	VX048785.D	09 Dec 2025 15:29	vial A pH#5.0 Need 10X	JC/MD	Dilution
20	Q3803-10	CHASE-J	VX048786.D	09 Dec 2025 15:49	Internal standard fail;Surrogate fail	JC/MD	ReRun
21	Q3803-11	CHASE-K	VX048787.D	09 Dec 2025 16:09	vial A pH#5.0 Surrogate fail	JC/MD	ReRun
22	Q3803-12	CHASE-L	VX048788.D	09 Dec 2025 16:30	vial A pH#5.0 Surrogate fail	JC/MD	ReRun
23	Q3803-13	CHASE-M	VX048789.D	09 Dec 2025 16:50	vial A pH#5.0 Surrogate fail	JC/MD	ReRun
24	Q3803-14	CHASE-N	VX048790.D	09 Dec 2025 17:11	vial A pH#5.0	JC/MD	Ok
25	IBLK	IBLK	VX048791.D	09 Dec 2025 17:31		JC/MD	Ok
26	VX1209WBSD02	VX1209WBSD02	VX048792.D	09 Dec 2025 17:52	Recovery fail	JC/MD	Not Ok
27	Q3788-02MS	OWBR-05-135-120525	VX048793.D	09 Dec 2025 18:12	vial A pH<2	JC/MD	Ok,M
28	VSTDCCC050	VSTDCCC050EC	VX048794.D	09 Dec 2025 18:32		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX121025

Review By	Mahesh Dadoda	Review On	12/11/2025 11:58:26 AM
Supervise By	Semsettin Yesilyurt	Supervise On	12/11/2025 12:34:32 PM
SubDirectory	VX121025	HP Acquire Method	HP Processing Method 82X120425W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136598
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136599,VP136600

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048795.D	10 Dec 2025 09:13		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX048796.D	10 Dec 2025 09:40	pH#lot#v14222	JC/MD	Ok,M
3	VX1210WBL01	VX1210WBL01	VX048797.D	10 Dec 2025 10:10		JC/MD	Ok
4	VX1210MBL01	VX1210MBL01	VX048798.D	10 Dec 2025 10:31		JC/MD	Ok
5	VX1210WBS01	VX1210WBS01	VX048799.D	10 Dec 2025 10:58		JC/MD	Ok,M
6	VX1210WBSD01	VX1210WBSD01	VX048800.D	10 Dec 2025 11:26		JC/MD	Ok,M
7	Q3803-08DL	CHASE-HDL	VX048801.D	10 Dec 2025 11:47	vial B pH#5.0	JC/MD	Ok
8	Q3803-09DL	CHASE-IDL	VX048802.D	10 Dec 2025 12:07	vial B pH#5.0	JC/MD	Ok
9	Q3803-11	CHASE-K	VX048803.D	10 Dec 2025 12:28	vial B pH#5.0	JC/MD	Ok
10	Q3803-12	CHASE-L	VX048804.D	10 Dec 2025 12:48	vial B pH#5.0	JC/MD	Ok,M
11	Q3803-13RE	CHASE-MRE	VX048805.D	10 Dec 2025 13:09	vial B pH#5.0 ,Surrogate Fail	JC/MD	Confirms
12	IBLK	IBLK	VX048806.D	10 Dec 2025 13:29		JC/MD	Ok
13	Q3787-07RE	OWBR-02-170-120425	VX048807.D	10 Dec 2025 13:49	vial B pH<2 ,Surrogate fail	JC/MD	Confirms
14	Q3787-09RE	OWBR-02-170-120425	VX048808.D	10 Dec 2025 14:10	vial B pH<2 ,Surrogate fail	JC/MD	Confirms
15	Q3787-11RE	OW-03B-51.5-120425	VX048809.D	10 Dec 2025 14:30	vial B pH<2 Surrogate fail	JC/MD	Confirms
16	Q3787-13RE	OW-03B-51.5-120425	VX048810.D	10 Dec 2025 14:51	vial B pH<2 Surrogate fail	JC/MD	Confirms
17	Q3787-17RE	OW-08B-72.5-120425	VX048811.D	10 Dec 2025 15:11	vial B pH<2 Surrogate fail	JC/MD	Confirms
18	Q3787-19	OW-02B-21.2-120425	VX048812.D	10 Dec 2025 15:32	vial B pH<2	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX121025

Review By	Maresh Dadoda	Review On	12/11/2025 11:58:26 AM
Supervise By	Semsettin Yesilyurt	Supervise On	12/11/2025 12:34:32 PM
SubDirectory	VX121025	HP Acquire Method	HP Processing Method 82X120425W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP136598		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136599,VP136600		

19	Q3787-21	TB01-120425	VX048813.D	10 Dec 2025 15:53	vial B pH<2 ,TB	JC/MD	Ok
20	Q3788-07	OW-01B-66.5-120525	VX048814.D	10 Dec 2025 16:13	vial B pH<2	JC/MD	Ok
21	Q3788-09RE	BR-05-465-120525RE	VX048815.D	10 Dec 2025 16:34	vial B pH<2 ,Surrogate fail	JC/MD	Confirms
22	Q3788-11RE	MW-16B-87.5-120525R	VX048816.D	10 Dec 2025 16:54	vial B pH<2 ,Surrogate fail	JC/MD	Confirms
23	Q3776-01	IDW-RAD-20251203	VX048817.D	10 Dec 2025 17:15	vial A pH<2 Surrogate fail	JC/MD	ReRun
24	Q3776-02	TB-20251203	VX048818.D	10 Dec 2025 17:35	vial A pH<2 ,TB;Surrogate fail	JC/MD	ReRun
25	PB170875TB	PB170875TB	VX048819.D	10 Dec 2025 17:56		JC/MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VX048820.D	10 Dec 2025 18:37		JC/MD	Ok,M

M : Manual Integration