

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JAC005
 Lab Code: ACE SDG NO.: Q3788
 Lab File ID: VX048768.D Date Analyzed: 12/09/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:25
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	159019	5.53	262535	6.74	233121	10.04
UPPER LIMIT	318038	6.032	525070	7.239	466242	10.537
LOWER LIMIT	79509.5	5.032	131268	6.239	116561	9.537
EPA SAMPLE NO.						
OWBR-05-135-120525	135193	5.54	279762	6.75	295744	10.04
OWBR-05-135-120525MS	135862	5.54	235253	6.75	212826	10.04
OWBR-05-135-120525MSD	135940	5.54	240144	6.75	217048	10.04
VX1209WBL01	129302	5.54	266304	6.75	288359	10.04
VX1209WBS02	172951	5.53	289026	6.74	260322	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.