

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JAC005
 Lab Code: ACE SDG NO.: Q3788
 Lab File ID: VX048796.D Date Analyzed: 12/10/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:40
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	164636	5.53	275167	6.74	243543	10.04
UPPER LIMIT	329272	6.031	550334	7.238	487086	10.537
LOWER LIMIT	82318	5.031	137584	6.238	121772	9.537
EPA SAMPLE NO.						
OW-01B-66.5-120525	134347	5.54	281654	6.75	303661	10.04
BR-05-465-120525RE	141844	5.54	294856	6.75	321925	10.04
MW-16B-87.5-120525RE	145186	5.54	297607	6.75	320506	10.04
VX1210WBL01	143182	5.53	288257	6.74	306883	10.04
VX1210WBS01	153473	5.53	261026	6.74	236663	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.