

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: JACO05

Lab Code: ACE

SDG NO.: Q3788

Lab File ID: VX048713.D

BFB Injection Date: 12/04/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:15

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.6
75	30.0 - 60.0% of mass 95	51.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	1 (1.3) 1
174	50.0 - 100.0% of mass 95	76.2
175	5.0 - 9.0% of mass 174	5.9 (7.8) 1
176	95.0 - 101.0% of mass 174	74.4 (97.7) 1
177	5.0 - 9.0% of mass 176	4.9 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX048714.D	12/04/2025	08:47
VSTDIC005	VSTDIC005	VX048715.D	12/04/2025	09:16
VSTDIC020	VSTDIC020	VX048716.D	12/04/2025	09:36
VSTDIC050	VSTDIC050	VX048717.D	12/04/2025	09:57
VSTDIC100	VSTDIC100	VX048718.D	12/04/2025	10:17
VSTDIC150	VSTDIC150	VX048719.D	12/04/2025	10:38