

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: JACO05

Lab Code: ACE

SDG NO.: Q3788

Lab File ID: VX048741.D

BFB Injection Date: 12/08/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:17

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.5
75	30.0 - 60.0% of mass 95	51.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.2
173	Less than 2.0% of mass 174	1 (1.2) 1
174	50.0 - 100.0% of mass 95	80.1
175	5.0 - 9.0% of mass 174	6 (7.5) 1
176	95.0 - 101.0% of mass 174	76.4 (95.3) 1
177	5.0 - 9.0% of mass 176	5.3 (7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX048742.D	12/08/2025	09:20
VX1208WBL01	VX1208WBL01	VX048744.D	12/08/2025	11:15
VX1208WBS01	VX1208WBS01	VX048745.D	12/08/2025	11:44
BR-05-465-120525	Q3788-09	VX048761.D	12/08/2025	17:21
MW-16B-87.5-120525	Q3788-11	VX048762.D	12/08/2025	17:41
TB01-120525	Q3788-13	VX048763.D	12/08/2025	18:02