

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Contract: TULL02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV147	Mercury	3.88	4.0	97	90 - 110	CV	12/09/2025	11:28	LB138165

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Contract: TULL02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV84	Mercury	5.20	5.0	104	90 - 110	CV	12/09/2025	11:33	LB138165
CCV85	Mercury	5.31	5.0	106	90 - 110	CV	12/09/2025	12:03	LB138165
CCV86	Mercury	5.28	5.0	106	90 - 110	CV	12/09/2025	12:40	LB138165
CCV87	Mercury	5.12	5.0	102	90 - 110	CV	12/09/2025	13:15	LB138165

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Contract: TULL02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7590	8000	95	90 - 110	P	12/09/2025	12:15	LB138174
	Antimony	3970	4000	99	90 - 110	P	12/09/2025	12:15	LB138174
	Arsenic	3860	4000	96	90 - 110	P	12/09/2025	12:15	LB138174
	Barium	7460	8000	93	90 - 110	P	12/09/2025	12:15	LB138174
	Beryllium	189	200	94	90 - 110	P	12/09/2025	12:15	LB138174
	Cadmium	1900	2000	95	90 - 110	P	12/09/2025	12:15	LB138174
	Calcium	19000	20000	95	90 - 110	P	12/09/2025	12:15	LB138174
	Chromium	788	800	98	90 - 110	P	12/09/2025	12:15	LB138174
	Cobalt	1890	2000	94	90 - 110	P	12/09/2025	12:15	LB138174
	Copper	975	1000	98	90 - 110	P	12/09/2025	12:15	LB138174
	Iron	3830	4000	96	90 - 110	P	12/09/2025	12:15	LB138174
	Lead	3750	4000	94	90 - 110	P	12/09/2025	12:15	LB138174
	Magnesium	19000	20000	95	90 - 110	P	12/09/2025	12:15	LB138174
	Manganese	1870	2000	94	90 - 110	P	12/09/2025	12:15	LB138174
	Nickel	1890	2000	94	90 - 110	P	12/09/2025	12:15	LB138174
	Potassium	19100	20000	96	90 - 110	P	12/09/2025	12:15	LB138174
	Selenium	3930	4000	98	90 - 110	P	12/09/2025	12:15	LB138174
	Silver	938	1000	94	90 - 110	P	12/09/2025	12:15	LB138174
	Sodium	18500	20000	92	90 - 110	P	12/09/2025	12:15	LB138174
	Thallium	3760	4000	94	90 - 110	P	12/09/2025	12:15	LB138174
	Vanadium	1930	2000	97	90 - 110	P	12/09/2025	12:15	LB138174
	Zinc	1980	2000	99	90 - 110	P	12/09/2025	12:15	LB138174

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Contract: TULL02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	99.7	100	100	80 - 120	P	12/09/2025	12:19	LB138174
	Antimony	55.1	50.0	110	80 - 120	P	12/09/2025	12:19	LB138174
	Arsenic	21.2	20.0	106	80 - 120	P	12/09/2025	12:19	LB138174
	Barium	99.5	100	100	80 - 120	P	12/09/2025	12:19	LB138174
	Beryllium	6.04	6.0	101	80 - 120	P	12/09/2025	12:19	LB138174
	Cadmium	5.98	6.0	100	80 - 120	P	12/09/2025	12:19	LB138174
	Calcium	2030	2000	102	80 - 120	P	12/09/2025	12:19	LB138174
	Chromium	10.0	10.0	100	80 - 120	P	12/09/2025	12:19	LB138174
	Cobalt	29.5	30.0	98	80 - 120	P	12/09/2025	12:19	LB138174
	Copper	21.1	20.0	106	80 - 120	P	12/09/2025	12:19	LB138174
	Iron	99.9	100	100	80 - 120	P	12/09/2025	12:19	LB138174
	Lead	10.6	12.0	88	80 - 120	P	12/09/2025	12:19	LB138174
	Magnesium	2100	2000	105	80 - 120	P	12/09/2025	12:19	LB138174
	Manganese	20.7	20.0	103	80 - 120	P	12/09/2025	12:19	LB138174
	Nickel	40.1	40.0	100	80 - 120	P	12/09/2025	12:19	LB138174
	Potassium	2050	2000	103	80 - 120	P	12/09/2025	12:19	LB138174
	Selenium	22.0	20.0	110	80 - 120	P	12/09/2025	12:19	LB138174
	Silver	10.4	10.0	104	80 - 120	P	12/09/2025	12:19	LB138174
	Sodium	2030	2000	101	80 - 120	P	12/09/2025	12:19	LB138174
	Thallium	43.8	40.0	110	80 - 120	P	12/09/2025	12:19	LB138174
	Vanadium	41.1	40.0	103	80 - 120	P	12/09/2025	12:19	LB138174
	Zinc	42.9	40.0	107	80 - 120	P	12/09/2025	12:19	LB138174

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Contract: TULL02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9680	10000	97	90 - 110	P	12/09/2025	12:49	LB138174
	Antimony	4920	5000	98	90 - 110	P	12/09/2025	12:49	LB138174
	Arsenic	4920	5000	98	90 - 110	P	12/09/2025	12:49	LB138174
	Barium	9670	10000	97	90 - 110	P	12/09/2025	12:49	LB138174
	Beryllium	237	250	95	90 - 110	P	12/09/2025	12:49	LB138174
	Cadmium	2430	2500	97	90 - 110	P	12/09/2025	12:49	LB138174
	Calcium	24200	25000	97	90 - 110	P	12/09/2025	12:49	LB138174
	Chromium	990	1000	99	90 - 110	P	12/09/2025	12:49	LB138174
	Cobalt	2430	2500	97	90 - 110	P	12/09/2025	12:49	LB138174
	Copper	1230	1250	99	90 - 110	P	12/09/2025	12:49	LB138174
	Iron	5110	5000	102	90 - 110	P	12/09/2025	12:49	LB138174
	Lead	4870	5000	97	90 - 110	P	12/09/2025	12:49	LB138174
	Magnesium	24200	25000	97	90 - 110	P	12/09/2025	12:49	LB138174
	Manganese	2430	2500	97	90 - 110	P	12/09/2025	12:49	LB138174
	Nickel	2430	2500	97	90 - 110	P	12/09/2025	12:49	LB138174
	Potassium	25200	25000	101	90 - 110	P	12/09/2025	12:49	LB138174
	Selenium	4950	5000	99	90 - 110	P	12/09/2025	12:49	LB138174
	Silver	1250	1250	100	90 - 110	P	12/09/2025	12:49	LB138174
	Sodium	25100	25000	100	90 - 110	P	12/09/2025	12:49	LB138174
	Thallium	5110	5000	102	90 - 110	P	12/09/2025	12:49	LB138174
	Vanadium	2430	2500	97	90 - 110	P	12/09/2025	12:49	LB138174
	Zinc	2510	2500	100	90 - 110	P	12/09/2025	12:49	LB138174
CCV02	Aluminum	9650	10000	96	90 - 110	P	12/09/2025	13:41	LB138174
	Antimony	5030	5000	101	90 - 110	P	12/09/2025	13:41	LB138174
	Arsenic	5030	5000	100	90 - 110	P	12/09/2025	13:41	LB138174
	Barium	9370	10000	94	90 - 110	P	12/09/2025	13:41	LB138174
	Beryllium	235	250	94	90 - 110	P	12/09/2025	13:41	LB138174
	Cadmium	2440	2500	98	90 - 110	P	12/09/2025	13:41	LB138174
	Calcium	23700	25000	95	90 - 110	P	12/09/2025	13:41	LB138174
	Chromium	981	1000	98	90 - 110	P	12/09/2025	13:41	LB138174
	Cobalt	2450	2500	98	90 - 110	P	12/09/2025	13:41	LB138174
	Copper	1250	1250	100	90 - 110	P	12/09/2025	13:41	LB138174
	Iron	4830	5000	96	90 - 110	P	12/09/2025	13:41	LB138174
	Lead	4900	5000	98	90 - 110	P	12/09/2025	13:41	LB138174

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Contract: TULL02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	23800	25000	95	90 - 110	P	12/09/2025	13:41	LB138174
	Manganese	2330	2500	93	90 - 110	P	12/09/2025	13:41	LB138174
	Nickel	2450	2500	98	90 - 110	P	12/09/2025	13:41	LB138174
	Potassium	24600	25000	98	90 - 110	P	12/09/2025	13:41	LB138174
	Selenium	5090	5000	102	90 - 110	P	12/09/2025	13:41	LB138174
	Silver	1230	1250	99	90 - 110	P	12/09/2025	13:41	LB138174
	Sodium	24400	25000	97	90 - 110	P	12/09/2025	13:41	LB138174
	Thallium	5160	5000	103	90 - 110	P	12/09/2025	13:41	LB138174
	Vanadium	2390	2500	96	90 - 110	P	12/09/2025	13:41	LB138174
	Zinc	2490	2500	100	90 - 110	P	12/09/2025	13:41	LB138174
CCV03	Aluminum	9660	10000	97	90 - 110	P	12/09/2025	15:09	LB138174
	Antimony	5000	5000	100	90 - 110	P	12/09/2025	15:09	LB138174
	Arsenic	4910	5000	98	90 - 110	P	12/09/2025	15:09	LB138174
	Barium	9800	10000	98	90 - 110	P	12/09/2025	15:09	LB138174
	Beryllium	226	250	90	90 - 110	P	12/09/2025	15:09	LB138174
	Cadmium	2410	2500	96	90 - 110	P	12/09/2025	15:09	LB138174
	Calcium	24200	25000	97	90 - 110	P	12/09/2025	15:09	LB138174
	Chromium	953	1000	95	90 - 110	P	12/09/2025	15:09	LB138174
	Cobalt	2430	2500	97	90 - 110	P	12/09/2025	15:09	LB138174
	Copper	1250	1250	100	90 - 110	P	12/09/2025	15:09	LB138174
	Iron	5110	5000	102	90 - 110	P	12/09/2025	15:09	LB138174
	Lead	4850	5000	97	90 - 110	P	12/09/2025	15:09	LB138174
	Magnesium	23800	25000	95	90 - 110	P	12/09/2025	15:09	LB138174
	Manganese	2450	2500	98	90 - 110	P	12/09/2025	15:09	LB138174
	Nickel	2430	2500	97	90 - 110	P	12/09/2025	15:09	LB138174
	Potassium	25300	25000	101	90 - 110	P	12/09/2025	15:09	LB138174
	Selenium	4990	5000	100	90 - 110	P	12/09/2025	15:09	LB138174
	Silver	1210	1250	97	90 - 110	P	12/09/2025	15:09	LB138174
	Sodium	25800	25000	103	90 - 110	P	12/09/2025	15:09	LB138174
	Thallium	4790	5000	96	90 - 110	P	12/09/2025	15:09	LB138174
	Vanadium	2440	2500	98	90 - 110	P	12/09/2025	15:09	LB138174
	Zinc	2410	2500	97	90 - 110	P	12/09/2025	15:09	LB138174
CCV04	Aluminum	9320	10000	93	90 - 110	P	12/09/2025	16:00	LB138174
	Antimony	4960	5000	99	90 - 110	P	12/09/2025	16:00	LB138174

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Contract: TULL02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	4910	5000	98	90 - 110	P	12/09/2025	16:00	LB138174
	Barium	9210	10000	92	90 - 110	P	12/09/2025	16:00	LB138174
	Beryllium	228	250	91	90 - 110	P	12/09/2025	16:00	LB138174
	Cadmium	2400	2500	96	90 - 110	P	12/09/2025	16:00	LB138174
	Calcium	23100	25000	92	90 - 110	P	12/09/2025	16:00	LB138174
	Chromium	993	1000	99	90 - 110	P	12/09/2025	16:00	LB138174
	Cobalt	2410	2500	97	90 - 110	P	12/09/2025	16:00	LB138174
	Copper	1240	1250	99	90 - 110	P	12/09/2025	16:00	LB138174
	Iron	4980	5000	100	90 - 110	P	12/09/2025	16:00	LB138174
	Lead	4820	5000	96	90 - 110	P	12/09/2025	16:00	LB138174
	Magnesium	22900	25000	92	90 - 110	P	12/09/2025	16:00	LB138174
	Manganese	2300	2500	92	90 - 110	P	12/09/2025	16:00	LB138174
	Nickel	2410	2500	96	90 - 110	P	12/09/2025	16:00	LB138174
	Potassium	25100	25000	100	90 - 110	P	12/09/2025	16:00	LB138174
	Selenium	4990	5000	100	90 - 110	P	12/09/2025	16:00	LB138174
	Silver	1260	1250	101	90 - 110	P	12/09/2025	16:00	LB138174
	Sodium	25200	25000	101	90 - 110	P	12/09/2025	16:00	LB138174
	Thallium	4920	5000	98	90 - 110	P	12/09/2025	16:00	LB138174
	Vanadium	2330	2500	93	90 - 110	P	12/09/2025	16:00	LB138174
	Zinc	2510	2500	100	90 - 110	P	12/09/2025	16:00	LB138174
CCV05	Aluminum	9810	10000	98	90 - 110	P	12/09/2025	16:50	LB138174
	Antimony	4970	5000	99	90 - 110	P	12/09/2025	16:50	LB138174
	Arsenic	4940	5000	99	90 - 110	P	12/09/2025	16:50	LB138174
	Barium	9410	10000	94	90 - 110	P	12/09/2025	16:50	LB138174
	Beryllium	229	250	92	90 - 110	P	12/09/2025	16:50	LB138174
	Cadmium	2420	2500	97	90 - 110	P	12/09/2025	16:50	LB138174
	Calcium	24200	25000	97	90 - 110	P	12/09/2025	16:50	LB138174
	Chromium	987	1000	99	90 - 110	P	12/09/2025	16:50	LB138174
	Cobalt	2430	2500	97	90 - 110	P	12/09/2025	16:50	LB138174
	Copper	1240	1250	99	90 - 110	P	12/09/2025	16:50	LB138174
	Iron	4980	5000	100	90 - 110	P	12/09/2025	16:50	LB138174
	Lead	4840	5000	97	90 - 110	P	12/09/2025	16:50	LB138174
	Magnesium	24200	25000	97	90 - 110	P	12/09/2025	16:50	LB138174
	Manganese	2370	2500	95	90 - 110	P	12/09/2025	16:50	LB138174

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Contract: TULL02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2420	2500	97	90 - 110	P	12/09/2025	16:50	LB138174
	Potassium	24700	25000	99	90 - 110	P	12/09/2025	16:50	LB138174
	Selenium	5000	5000	100	90 - 110	P	12/09/2025	16:50	LB138174
	Silver	1210	1250	97	90 - 110	P	12/09/2025	16:50	LB138174
	Sodium	24200	25000	97	90 - 110	P	12/09/2025	16:50	LB138174
	Thallium	4980	5000	100	90 - 110	P	12/09/2025	16:50	LB138174
	Vanadium	2450	2500	98	90 - 110	P	12/09/2025	16:50	LB138174
	Zinc	2450	2500	98	90 - 110	P	12/09/2025	16:50	LB138174
CCV06	Aluminum	9660	10000	97	90 - 110	P	12/09/2025	17:41	LB138174
	Antimony	5010	5000	100	90 - 110	P	12/09/2025	17:41	LB138174
	Arsenic	4970	5000	99	90 - 110	P	12/09/2025	17:41	LB138174
	Barium	9340	10000	93	90 - 110	P	12/09/2025	17:41	LB138174
	Beryllium	232	250	93	90 - 110	P	12/09/2025	17:41	LB138174
	Cadmium	2410	2500	97	90 - 110	P	12/09/2025	17:41	LB138174
	Calcium	23700	25000	95	90 - 110	P	12/09/2025	17:41	LB138174
	Chromium	986	1000	99	90 - 110	P	12/09/2025	17:41	LB138174
	Cobalt	2430	2500	97	90 - 110	P	12/09/2025	17:41	LB138174
	Copper	1250	1250	100	90 - 110	P	12/09/2025	17:41	LB138174
	Iron	4900	5000	98	90 - 110	P	12/09/2025	17:41	LB138174
	Lead	4840	5000	97	90 - 110	P	12/09/2025	17:41	LB138174
	Magnesium	23700	25000	95	90 - 110	P	12/09/2025	17:41	LB138174
	Manganese	2320	2500	93	90 - 110	P	12/09/2025	17:41	LB138174
	Nickel	2420	2500	97	90 - 110	P	12/09/2025	17:41	LB138174
	Potassium	24800	25000	99	90 - 110	P	12/09/2025	17:41	LB138174
	Selenium	5050	5000	101	90 - 110	P	12/09/2025	17:41	LB138174
	Silver	1240	1250	99	90 - 110	P	12/09/2025	17:41	LB138174
	Sodium	24400	25000	98	90 - 110	P	12/09/2025	17:41	LB138174
	Thallium	4800	5000	96	90 - 110	P	12/09/2025	17:41	LB138174
	Vanadium	2400	2500	96	90 - 110	P	12/09/2025	17:41	LB138174
	Zinc	2490	2500	100	90 - 110	P	12/09/2025	17:41	LB138174
CCV07	Aluminum	9570	10000	96	90 - 110	P	12/09/2025	18:31	LB138174
	Antimony	4890	5000	98	90 - 110	P	12/09/2025	18:31	LB138174
	Arsenic	4800	5000	96	90 - 110	P	12/09/2025	18:31	LB138174
	Barium	9690	10000	97	90 - 110	P	12/09/2025	18:31	LB138174

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Contract: TULL02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	227	250	91	90 - 110	P	12/09/2025	18:31	LB138174
	Cadmium	2350	2500	94	90 - 110	P	12/09/2025	18:31	LB138174
	Calcium	23800	25000	95	90 - 110	P	12/09/2025	18:31	LB138174
	Chromium	951	1000	95	90 - 110	P	12/09/2025	18:31	LB138174
	Cobalt	2380	2500	95	90 - 110	P	12/09/2025	18:31	LB138174
	Copper	1220	1250	98	90 - 110	P	12/09/2025	18:31	LB138174
	Iron	5080	5000	102	90 - 110	P	12/09/2025	18:31	LB138174
	Lead	4730	5000	95	90 - 110	P	12/09/2025	18:31	LB138174
	Magnesium	23500	25000	94	90 - 110	P	12/09/2025	18:31	LB138174
	Manganese	2410	2500	96	90 - 110	P	12/09/2025	18:31	LB138174
	Nickel	2370	2500	95	90 - 110	P	12/09/2025	18:31	LB138174
	Potassium	25100	25000	100	90 - 110	P	12/09/2025	18:31	LB138174
	Selenium	4870	5000	97	90 - 110	P	12/09/2025	18:31	LB138174
	Silver	1220	1250	98	90 - 110	P	12/09/2025	18:31	LB138174
	Sodium	25500	25000	102	90 - 110	P	12/09/2025	18:31	LB138174
	Thallium	4650	5000	93	90 - 110	P	12/09/2025	18:31	LB138174
	Vanadium	2400	2500	96	90 - 110	P	12/09/2025	18:31	LB138174
	Zinc	2460	2500	98	90 - 110	P	12/09/2025	18:31	LB138174
CCV08	Aluminum	9450	10000	94	90 - 110	P	12/09/2025	19:30	LB138174
	Antimony	4850	5000	97	90 - 110	P	12/09/2025	19:30	LB138174
	Arsenic	4790	5000	96	90 - 110	P	12/09/2025	19:30	LB138174
	Barium	9180	10000	92	90 - 110	P	12/09/2025	19:30	LB138174
	Beryllium	225	250	90	90 - 110	P	12/09/2025	19:30	LB138174
	Cadmium	2350	2500	94	90 - 110	P	12/09/2025	19:30	LB138174
	Calcium	23400	25000	94	90 - 110	P	12/09/2025	19:30	LB138174
	Chromium	956	1000	96	90 - 110	P	12/09/2025	19:30	LB138174
	Cobalt	2370	2500	95	90 - 110	P	12/09/2025	19:30	LB138174
	Copper	1210	1250	97	90 - 110	P	12/09/2025	19:30	LB138174
	Iron	4910	5000	98	90 - 110	P	12/09/2025	19:30	LB138174
	Lead	4740	5000	95	90 - 110	P	12/09/2025	19:30	LB138174
	Magnesium	23300	25000	93	90 - 110	P	12/09/2025	19:30	LB138174
	Manganese	2320	2500	93	90 - 110	P	12/09/2025	19:30	LB138174
	Nickel	2360	2500	95	90 - 110	P	12/09/2025	19:30	LB138174
	Potassium	24300	25000	97	90 - 110	P	12/09/2025	19:30	LB138174

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Contract: TULL02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	4840	5000	97	90 - 110	P	12/09/2025	19:30	LB138174
	Silver	1200	1250	96	90 - 110	P	12/09/2025	19:30	LB138174
	Sodium	24100	25000	96	90 - 110	P	12/09/2025	19:30	LB138174
	Thallium	4690	5000	94	90 - 110	P	12/09/2025	19:30	LB138174
	Vanadium	2360	2500	94	90 - 110	P	12/09/2025	19:30	LB138174
	Zinc	2410	2500	96	90 - 110	P	12/09/2025	19:30	LB138174
CCV09	Aluminum	9360	10000	94	90 - 110	P	12/09/2025	20:37	LB138174
	Antimony	4880	5000	98	90 - 110	P	12/09/2025	20:37	LB138174
	Arsenic	4800	5000	96	90 - 110	P	12/09/2025	20:37	LB138174
	Barium	9610	10000	96	90 - 110	P	12/09/2025	20:37	LB138174
	Beryllium	241	250	96	90 - 110	P	12/09/2025	20:37	LB138174
	Cadmium	2350	2500	94	90 - 110	P	12/09/2025	20:37	LB138174
	Calcium	23200	25000	93	90 - 110	P	12/09/2025	20:37	LB138174
	Chromium	962	1000	96	90 - 110	P	12/09/2025	20:37	LB138174
	Cobalt	2380	2500	95	90 - 110	P	12/09/2025	20:37	LB138174
	Copper	1220	1250	98	90 - 110	P	12/09/2025	20:37	LB138174
	Iron	5130	5000	103	90 - 110	P	12/09/2025	20:37	LB138174
	Lead	4730	5000	95	90 - 110	P	12/09/2025	20:37	LB138174
	Magnesium	22900	25000	92	90 - 110	P	12/09/2025	20:37	LB138174
	Manganese	2320	2500	93	90 - 110	P	12/09/2025	20:37	LB138174
	Nickel	2370	2500	95	90 - 110	P	12/09/2025	20:37	LB138174
	Potassium	25300	25000	101	90 - 110	P	12/09/2025	20:37	LB138174
	Selenium	4850	5000	97	90 - 110	P	12/09/2025	20:37	LB138174
	Silver	1220	1250	98	90 - 110	P	12/09/2025	20:37	LB138174
	Sodium	25200	25000	101	90 - 110	P	12/09/2025	20:37	LB138174
	Thallium	5060	5000	101	90 - 110	P	12/09/2025	20:37	LB138174
	Vanadium	2360	2500	94	90 - 110	P	12/09/2025	20:37	LB138174
	Zinc	2460	2500	98	90 - 110	P	12/09/2025	20:37	LB138174
CCV10	Aluminum	9390	10000	94	90 - 110	P	12/09/2025	21:27	LB138174
	Antimony	4960	5000	99	90 - 110	P	12/09/2025	21:27	LB138174
	Arsenic	4890	5000	98	90 - 110	P	12/09/2025	21:27	LB138174
	Barium	9460	10000	95	90 - 110	P	12/09/2025	21:27	LB138174
	Beryllium	239	250	96	90 - 110	P	12/09/2025	21:27	LB138174
	Cadmium	2360	2500	94	90 - 110	P	12/09/2025	21:27	LB138174

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Contract: TULL02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Calcium	23100	25000	92	90 - 110	P	12/09/2025	21:27	LB138174
	Chromium	948	1000	95	90 - 110	P	12/09/2025	21:27	LB138174
	Cobalt	2380	2500	95	90 - 110	P	12/09/2025	21:27	LB138174
	Copper	1230	1250	98	90 - 110	P	12/09/2025	21:27	LB138174
	Iron	4950	5000	99	90 - 110	P	12/09/2025	21:27	LB138174
	Lead	4760	5000	95	90 - 110	P	12/09/2025	21:27	LB138174
	Magnesium	22900	25000	91	90 - 110	P	12/09/2025	21:27	LB138174
	Manganese	2280	2500	91	90 - 110	P	12/09/2025	21:27	LB138174
	Nickel	2370	2500	95	90 - 110	P	12/09/2025	21:27	LB138174
	Potassium	24600	25000	98	90 - 110	P	12/09/2025	21:27	LB138174
	Selenium	4990	5000	100	90 - 110	P	12/09/2025	21:27	LB138174
	Silver	1190	1250	95	90 - 110	P	12/09/2025	21:27	LB138174
	Sodium	24400	25000	98	90 - 110	P	12/09/2025	21:27	LB138174
	Thallium	5230	5000	105	90 - 110	P	12/09/2025	21:27	LB138174
	Vanadium	2330	2500	93	90 - 110	P	12/09/2025	21:27	LB138174
	Zinc	2430	2500	97	90 - 110	P	12/09/2025	21:27	LB138174
CCV11	Aluminum	9410	10000	94	90 - 110	P	12/09/2025	22:17	LB138174
	Antimony	4930	5000	99	90 - 110	P	12/09/2025	22:17	LB138174
	Arsenic	4850	5000	97	90 - 110	P	12/09/2025	22:17	LB138174
	Barium	9600	10000	96	90 - 110	P	12/09/2025	22:17	LB138174
	Beryllium	248	250	99	90 - 110	P	12/09/2025	22:17	LB138174
	Cadmium	2370	2500	95	90 - 110	P	12/09/2025	22:17	LB138174
	Calcium	23200	25000	93	90 - 110	P	12/09/2025	22:17	LB138174
	Chromium	963	1000	96	90 - 110	P	12/09/2025	22:17	LB138174
	Cobalt	2400	2500	96	90 - 110	P	12/09/2025	22:17	LB138174
	Copper	1230	1250	98	90 - 110	P	12/09/2025	22:17	LB138174
	Iron	4940	5000	99	90 - 110	P	12/09/2025	22:17	LB138174
	Lead	4760	5000	95	90 - 110	P	12/09/2025	22:17	LB138174
	Magnesium	23000	25000	92	90 - 110	P	12/09/2025	22:17	LB138174
	Manganese	2280	2500	91	90 - 110	P	12/09/2025	22:17	LB138174
	Nickel	2390	2500	95	90 - 110	P	12/09/2025	22:17	LB138174
	Potassium	24500	25000	98	90 - 110	P	12/09/2025	22:17	LB138174
	Selenium	4920	5000	98	90 - 110	P	12/09/2025	22:17	LB138174
	Silver	1190	1250	95	90 - 110	P	12/09/2025	22:17	LB138174

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Contract: TULL02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV11	Sodium	24100	25000	96	90 - 110	P	12/09/2025	22:17	LB138174
	Thallium	5240	5000	105	90 - 110	P	12/09/2025	22:17	LB138174
	Vanadium	2350	2500	94	90 - 110	P	12/09/2025	22:17	LB138174
	Zinc	2440	2500	97	90 - 110	P	12/09/2025	22:17	LB138174
CCV12	Aluminum	9590	10000	96	90 - 110	P	12/09/2025	23:20	LB138174
	Antimony	4970	5000	100	90 - 110	P	12/09/2025	23:20	LB138174
	Arsenic	4830	5000	97	90 - 110	P	12/09/2025	23:20	LB138174
	Barium	9960	10000	100	90 - 110	P	12/09/2025	23:20	LB138174
	Beryllium	251	250	100	90 - 110	P	12/09/2025	23:20	LB138174
	Cadmium	2370	2500	95	90 - 110	P	12/09/2025	23:20	LB138174
	Calcium	23700	25000	95	90 - 110	P	12/09/2025	23:20	LB138174
	Chromium	943	1000	94	90 - 110	P	12/09/2025	23:20	LB138174
	Cobalt	2410	2500	96	90 - 110	P	12/09/2025	23:20	LB138174
	Copper	1240	1250	99	90 - 110	P	12/09/2025	23:20	LB138174
	Iron	5040	5000	101	90 - 110	P	12/09/2025	23:20	LB138174
	Lead	4790	5000	96	90 - 110	P	12/09/2025	23:20	LB138174
	Magnesium	23400	25000	94	90 - 110	P	12/09/2025	23:20	LB138174
	Manganese	2400	2500	96	90 - 110	P	12/09/2025	23:20	LB138174
	Nickel	2400	2500	96	90 - 110	P	12/09/2025	23:20	LB138174
	Potassium	25000	25000	100	90 - 110	P	12/09/2025	23:20	LB138174
	Selenium	4920	5000	98	90 - 110	P	12/09/2025	23:20	LB138174
	Silver	1220	1250	98	90 - 110	P	12/09/2025	23:20	LB138174
	Sodium	25500	25000	102	90 - 110	P	12/09/2025	23:20	LB138174
	Thallium	5030	5000	101	90 - 110	P	12/09/2025	23:20	LB138174
	Vanadium	2400	2500	96	90 - 110	P	12/09/2025	23:20	LB138174
	Zinc	2470	2500	99	90 - 110	P	12/09/2025	23:20	LB138174
CCV13	Aluminum	9730	10000	97	90 - 110	P	12/09/2025	23:41	LB138174
	Antimony	5110	5000	102	90 - 110	P	12/09/2025	23:41	LB138174
	Arsenic	5020	5000	100	90 - 110	P	12/09/2025	23:41	LB138174
	Barium	9690	10000	97	90 - 110	P	12/09/2025	23:41	LB138174
	Beryllium	252	250	101	90 - 110	P	12/09/2025	23:41	LB138174
	Cadmium	2440	2500	97	90 - 110	P	12/09/2025	23:41	LB138174
	Calcium	23800	25000	95	90 - 110	P	12/09/2025	23:41	LB138174
	Chromium	966	1000	97	90 - 110	P	12/09/2025	23:41	LB138174

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Contract: TULL02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV13	Cobalt	2470	2500	99	90 - 110	P	12/09/2025	23:41	LB138174
	Copper	1270	1250	102	90 - 110	P	12/09/2025	23:41	LB138174
	Iron	4960	5000	99	90 - 110	P	12/09/2025	23:41	LB138174
	Lead	4900	5000	98	90 - 110	P	12/09/2025	23:41	LB138174
	Magnesium	23700	25000	95	90 - 110	P	12/09/2025	23:41	LB138174
	Manganese	2350	2500	94	90 - 110	P	12/09/2025	23:41	LB138174
	Nickel	2450	2500	98	90 - 110	P	12/09/2025	23:41	LB138174
	Potassium	24800	25000	99	90 - 110	P	12/09/2025	23:41	LB138174
	Selenium	5120	5000	102	90 - 110	P	12/09/2025	23:41	LB138174
	Silver	1220	1250	98	90 - 110	P	12/09/2025	23:41	LB138174
	Sodium	24800	25000	99	90 - 110	P	12/09/2025	23:41	LB138174
	Thallium	5350	5000	107	90 - 110	P	12/09/2025	23:41	LB138174
	Vanadium	2420	2500	97	90 - 110	P	12/09/2025	23:41	LB138174
	Zinc	2490	2500	100	90 - 110	P	12/09/2025	23:41	LB138174