



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Contract: TULL02

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.17	0.2	83	70 - 130	CV	12/09/2025	11:40	LB138165
CRI01	Aluminum	104	100	104	65 - 135	P	12/09/2025	12:28	LB138174
	Antimony	54.0	50.0	108	65 - 135	P	12/09/2025	12:28	LB138174
	Arsenic	19.1	20.0	96	65 - 135	P	12/09/2025	12:28	LB138174
	Barium	96.1	100	96	65 - 135	P	12/09/2025	12:28	LB138174
	Beryllium	6.05	6.0	101	65 - 135	P	12/09/2025	12:28	LB138174
	Cadmium	5.84	6.0	97	65 - 135	P	12/09/2025	12:28	LB138174
	Calcium	2020	2000	101	65 - 135	P	12/09/2025	12:28	LB138174
	Chromium	10.4	10.0	104	65 - 135	P	12/09/2025	12:28	LB138174
	Cobalt	29.3	30.0	98	65 - 135	P	12/09/2025	12:28	LB138174
	Copper	21.5	20.0	107	65 - 135	P	12/09/2025	12:28	LB138174
	Iron	86.6	100	87	65 - 135	P	12/09/2025	12:28	LB138174
	Lead	12.0	12.0	100	65 - 135	P	12/09/2025	12:28	LB138174
	Magnesium	2090	2000	104	65 - 135	P	12/09/2025	12:28	LB138174
	Manganese	19.8	20.0	99	65 - 135	P	12/09/2025	12:28	LB138174
	Nickel	39.7	40.0	99	65 - 135	P	12/09/2025	12:28	LB138174
	Potassium	1970	2000	98	65 - 135	P	12/09/2025	12:28	LB138174
	Selenium	19.3	20.0	97	65 - 135	P	12/09/2025	12:28	LB138174
	Silver	10.0	10.0	100	65 - 135	P	12/09/2025	12:28	LB138174
	Sodium	1860	2000	93	65 - 135	P	12/09/2025	12:28	LB138174
	Thallium	44.4	40.0	111	65 - 135	P	12/09/2025	12:28	LB138174
	Vanadium	38.4	40.0	96	65 - 135	P	12/09/2025	12:28	LB138174
	Zinc	41.9	40.0	105	65 - 135	P	12/09/2025	12:28	LB138174