

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Contract: TULL02

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	239000	255000	94	216000	294000	12/09/2025	12:32	LB138174
	Antimony	2.17			-50	50	12/09/2025	12:32	LB138174
	Arsenic	5.20			-20	20	12/09/2025	12:32	LB138174
	Barium	5.08	6.0	85	-94	106	12/09/2025	12:32	LB138174
	Beryllium	1.40			-6	6	12/09/2025	12:32	LB138174
	Cadmium	4.58	1.0	458	-5	7	12/09/2025	12:32	LB138174
	Calcium	225000	245000	92	208000	282000	12/09/2025	12:32	LB138174
	Chromium	45.9	52.0	88	42	62	12/09/2025	12:32	LB138174
	Cobalt	2.34			-30	30	12/09/2025	12:32	LB138174
	Copper	-2.36	2.0	118	-18	22	12/09/2025	12:32	LB138174
	Iron	94200	101000	93	85600	116500	12/09/2025	12:32	LB138174
	Lead	3.51			-12	12	12/09/2025	12:32	LB138174
	Magnesium	243000	255000	95	216000	294000	12/09/2025	12:32	LB138174
	Manganese	8.08	7.0	115	-13	27	12/09/2025	12:32	LB138174
	Nickel	6.00	2.0	300	-38	42	12/09/2025	12:32	LB138174
	Potassium	101			0	0	12/09/2025	12:32	LB138174
	Selenium	1.24			-20	20	12/09/2025	12:32	LB138174
	Silver	-1.30			-10	10	12/09/2025	12:32	LB138174
	Sodium	64.9			0	0	12/09/2025	12:32	LB138174
	Thallium	28.4			-40	40	12/09/2025	12:32	LB138174
	Vanadium	5.43			-40	40	12/09/2025	12:32	LB138174
	Zinc	5.79			-40	40	12/09/2025	12:32	LB138174
ICSAB01	Aluminum	239000	247000	97	209000	285000	12/09/2025	12:37	LB138174
	Antimony	595	618	96	525	711	12/09/2025	12:37	LB138174
	Arsenic	107	104	103	88.4	120	12/09/2025	12:37	LB138174
	Barium	456	537	85	437	637	12/09/2025	12:37	LB138174
	Beryllium	468	495	94	420	570	12/09/2025	12:37	LB138174
	Cadmium	979	972	101	826	1120	12/09/2025	12:37	LB138174
	Calcium	225000	235000	96	199000	271000	12/09/2025	12:37	LB138174
	Chromium	545	542	101	460	624	12/09/2025	12:37	LB138174
	Cobalt	487	476	102	404	548	12/09/2025	12:37	LB138174
	Copper	464	511	91	434	588	12/09/2025	12:37	LB138174
	Iron	95200	99300	96	84400	114500	12/09/2025	12:37	LB138174
	Lead	53.9	49.0	110	37	61	12/09/2025	12:37	LB138174
	Magnesium	243000	248000	98	210000	286000	12/09/2025	12:37	LB138174
	Manganese	457	507	90	430	584	12/09/2025	12:37	LB138174
	Nickel	965	954	101	810	1100	12/09/2025	12:37	LB138174
	Potassium	42.5			0	0	12/09/2025	12:37	LB138174
	Selenium	49.8	46.0	108	26	66	12/09/2025	12:37	LB138174
	Silver	180	201	90	170	232	12/09/2025	12:37	LB138174
	Sodium	87.5			0	0	12/09/2025	12:37	LB138174
	Thallium	148	108	137	68	148	12/09/2025	12:37	LB138174

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Contract: TULL02

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	465	491	95	417	565	12/09/2025	12:37	LB138174
	Zinc	1040	952	109	809	1095	12/09/2025	12:37	LB138174