

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Contract: TULL02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170856BS							
Aluminum	mg/Kg	100	94.8		95	80 - 120	P
Antimony	mg/Kg	40.0	40.8		102	80 - 120	P
Arsenic	mg/Kg	40.0	39.4		98	80 - 120	P
Barium	mg/Kg	10.0	9.18		92	80 - 120	P
Beryllium	mg/Kg	10.0	9.17		92	80 - 120	P
Cadmium	mg/Kg	10.0	9.52		95	80 - 120	P
Calcium	mg/Kg	50.0	48.5	J	97	80 - 120	P
Chromium	mg/Kg	20.0	19.7		98	80 - 120	P
Cobalt	mg/Kg	10.0	9.59		96	80 - 120	P
Copper	mg/Kg	15.0	15.4		103	80 - 120	P
Iron	mg/Kg	150	147		98	80 - 120	P
Lead	mg/Kg	50.0	47.0		94	80 - 120	P
Magnesium	mg/Kg	100	94.7	J	95	80 - 120	P
Manganese	mg/Kg	10.0	9.41		94	80 - 120	P
Nickel	mg/Kg	25.0	24.1		96	80 - 120	P
Potassium	mg/Kg	500	478		96	80 - 120	P
Selenium	mg/Kg	100	102		102	80 - 120	P
Silver	mg/Kg	3.8	3.70		97	80 - 120	P
Sodium	mg/Kg	150	143		95	80 - 120	P
Thallium	mg/Kg	100	94.3		94	80 - 120	P
Vanadium	mg/Kg	15.0	14.3		95	80 - 120	P
Zinc	mg/Kg	10.0	10.4		104	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Contract: TULL02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170867BS Mercury	mg/Kg	0.26	0.21		82	80 - 120	CV