



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Contract: TULL02

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.18	0.2	91	70 - 130	CV	12/10/2025	09:22	LB138176
CRI01	Arsenic	20.0	20.0	100	65 - 135	P	12/10/2025	13:33	LB138187
	Barium	97.5	100	98	65 - 135	P	12/10/2025	13:33	LB138187
	Cadmium	5.99	6.0	100	65 - 135	P	12/10/2025	13:33	LB138187
	Chromium	9.47	10.0	95	65 - 135	P	12/10/2025	13:33	LB138187
	Lead	13.7	12.0	114	65 - 135	P	12/10/2025	13:33	LB138187
	Selenium	18.3	20.0	92	65 - 135	P	12/10/2025	13:33	LB138187
	Silver	9.55	10.0	96	65 - 135	P	12/10/2025	13:33	LB138187