

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170857TB		WATER		Batch Number:	PB170871		Prep Date:	12/09/2025	
	Mercury	0.76	<2	U	2.00	CV	12/10/2025	10:38	LB138176
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170871BL		WATER		Batch Number:	PB170871		Prep Date:	12/09/2025	
	Mercury	0.076	<0.2	U	0.20	CV	12/10/2025	09:29	LB138176

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170857TB		WATER		Batch Number:	PB170887		Prep Date:	12/10/2025	
	Arsenic	25.6	<50	U	100	P	12/10/2025	19:55	LB138187
	Barium	72.8	<250	U	500	P	12/10/2025	19:55	LB138187
	Cadmium	2.50	<15	U	30.0	P	12/10/2025	19:55	LB138187
	Chromium	21.1	<25	J	50.0	P	12/10/2025	19:55	LB138187
	Lead	17.0	<30	J	60.0	P	12/10/2025	19:55	LB138187
	Selenium	48.2	<50	U	100	P	12/10/2025	19:55	LB138187
	Silver	8.10	<25	U	50.0	P	12/10/2025	19:55	LB138187
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170887BL		WATER		Batch Number:	PB170887		Prep Date:	12/10/2025	
	Arsenic	25.6	<50	U	100	P	12/10/2025	20:04	LB138187
	Barium	72.8	<250	U	500	P	12/10/2025	20:04	LB138187
	Cadmium	2.50	<15	U	30.0	P	12/10/2025	20:04	LB138187
	Chromium	10.6	<25	U	50.0	P	12/10/2025	20:04	LB138187
	Lead	11.5	<30	U	60.0	P	12/10/2025	20:04	LB138187
	Selenium	48.2	<50	U	100	P	12/10/2025	20:04	LB138187
	Silver	8.10	<25	U	50.0	P	12/10/2025	20:04	LB138187