

**Metals**  
**- 4 -**  
**INTERFERENCE CHECK SAMPLE**

**Client:** Tully Construction Co., Inc.

**SDG No.:** Q3790

**Contract:** TULL02

**Lab Code:** ACE

**ICS Source:** EPA

**Instrument ID:** P4

| Sample ID      | Analyte  | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Low<br>Limit<br>(ug/L) | High<br>Limit<br>(ug/L) | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|----------------|----------|----------------|--------------------|---------------|------------------------|-------------------------|------------------|------------------|---------------|
| <b>ICSA01</b>  | Arsenic  | 7.31           |                    |               | -20                    | 20                      | 12/10/2025       | 13:38            | LB138187      |
|                | Barium   | 3.06           | 6.0                | 51            | -94                    | 106                     | 12/10/2025       | 13:38            | LB138187      |
|                | Cadmium  | 2.15           | 1.0                | 215           | -5                     | 7                       | 12/10/2025       | 13:38            | LB138187      |
|                | Chromium | 42.5           | 52.0               | 82            | 42                     | 62                      | 12/10/2025       | 13:38            | LB138187      |
|                | Lead     | 8.91           |                    |               | -12                    | 12                      | 12/10/2025       | 13:38            | LB138187      |
|                | Selenium | -1.91          |                    |               | -20                    | 20                      | 12/10/2025       | 13:38            | LB138187      |
|                | Silver   | 5.22           |                    |               | -10                    | 10                      | 12/10/2025       | 13:38            | LB138187      |
| <b>ICSAB01</b> | Arsenic  | 99.4           | 104                | 96            | 88.4                   | 120                     | 12/10/2025       | 13:42            | LB138187      |
|                | Barium   | 496            | 537                | 92            | 437                    | 637                     | 12/10/2025       | 13:42            | LB138187      |
|                | Cadmium  | 941            | 972                | 97            | 826                    | 1120                    | 12/10/2025       | 13:42            | LB138187      |
|                | Chromium | 502            | 542                | 93            | 460                    | 624                     | 12/10/2025       | 13:42            | LB138187      |
|                | Lead     | 59.8           | 49.0               | 122           | 37                     | 61                      | 12/10/2025       | 13:42            | LB138187      |
|                | Selenium | 42.4           | 46.0               | 92            | 26                     | 66                      | 12/10/2025       | 13:42            | LB138187      |
|                | Silver   | 178            | 201                | 89            | 170                    | 232                     | 12/10/2025       | 13:42            | LB138187      |