

**Metals**

- 7 -

**LABORATORY CONTROL SAMPLE SUMMARY**

**Client:** Tully Construction Co., Inc.

**SDG No.:** Q3790

**Contract:** TULL02

**Lab Code:** ACE

| Analyte               | Units | True Value | Result | C | %<br>Recovery | Acceptance<br>Limits | M  |
|-----------------------|-------|------------|--------|---|---------------|----------------------|----|
| PB170871BS<br>Mercury | ug/L  | 4.0        | 3.88   |   | 97            | 80 - 120             | CV |

**Metals**

- 7 -

**LABORATORY CONTROL SAMPLE SUMMARY**

**Client:** Tully Construction Co., Inc.

**SDG No.:** Q3790

**Contract:** TULL02

**Lab Code:** ACE

| Analyte    | Units | True Value | Result | C | %<br>Recovery | Acceptance<br>Limits | M |
|------------|-------|------------|--------|---|---------------|----------------------|---|
| PB170887BS |       |            |        |   |               |                      |   |
| Arsenic    | ug/L  | 4000       | 3800   |   | 95            | 80 - 120             | P |
| Barium     | ug/L  | 1000       | 918    |   | 92            | 80 - 120             | P |
| Cadmium    | ug/L  | 1000       | 924    |   | 92            | 80 - 120             | P |
| Chromium   | ug/L  | 2000       | 1940   |   | 97            | 80 - 120             | P |
| Lead       | ug/L  | 5000       | 4600   |   | 92            | 80 - 120             | P |
| Selenium   | ug/L  | 10000      | 9720   |   | 97            | 80 - 120             | P |
| Silver     | ug/L  | 380        | 355    |   | 93            | 80 - 120             | P |