

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TULL02

Lab Code: ACE

SDG NO.: Q3790

Continuing Calib Date: 12/09/2025

Initial Calibration Date(s): 11/21/2025

11/21/2025

Continuing Calib Time: 04:19

Initial Calibration Time(s): 10:44

19:39

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.76       | 5.77      | 5.67      | 5.87  | 0.01       |
| Aroclor-1016-2       | (2) | 5.79       | 5.79      | 5.69      | 5.89  | 0.01       |
| Aroclor-1016-3       | (3) | 5.85       | 5.85      | 5.75      | 5.95  | 0.00       |
| Aroclor-1016-4       | (4) | 5.94       | 5.95      | 5.85      | 6.05  | 0.01       |
| Aroclor-1016-5       | (5) | 6.24       | 6.24      | 6.14      | 6.34  | 0.00       |
| Aroclor-1260-1       | (1) | 7.35       | 7.36      | 7.26      | 7.46  | 0.01       |
| Aroclor-1260-2       | (2) | 7.61       | 7.61      | 7.51      | 7.71  | 0.00       |
| Aroclor-1260-3       | (3) | 7.96       | 7.97      | 7.87      | 8.07  | 0.01       |
| Aroclor-1260-4       | (4) | 8.19       | 8.19      | 8.09      | 8.29  | 0.00       |
| Aroclor-1260-5       | (5) | 8.51       | 8.52      | 8.42      | 8.62  | 0.01       |
| Tetrachloro-m-xylene |     | 4.61       | 4.61      | 4.51      | 4.71  | 0.00       |
| Decachlorobiphenyl   |     | 10.36      | 10.37     | 10.27     | 10.47 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TULL02

Lab Code: ACE

SDG NO.: Q3790

Continuing Calib Date: 12/09/2025

Initial Calibration Date(s): 11/21/2025

11/21/2025

Continuing Calib Time: 04:19

Initial Calibration Time(s): 10:44

19:39

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.84       | 4.85      | 4.75      | 4.95 | 0.01       |
| Aroclor-1016-2       | (2) | 4.86       | 4.87      | 4.77      | 4.97 | 0.01       |
| Aroclor-1016-3       | (3) | 5.04       | 5.04      | 4.94      | 5.14 | 0.00       |
| Aroclor-1016-4       | (4) | 5.08       | 5.08      | 4.98      | 5.18 | 0.00       |
| Aroclor-1016-5       | (5) | 5.29       | 5.30      | 5.20      | 5.40 | 0.01       |
| Aroclor-1260-1       | (1) | 6.32       | 6.32      | 6.22      | 6.42 | 0.00       |
| Aroclor-1260-2       | (2) | 6.51       | 6.51      | 6.41      | 6.61 | 0.00       |
| Aroclor-1260-3       | (3) | 6.66       | 6.66      | 6.56      | 6.76 | 0.00       |
| Aroclor-1260-4       | (4) | 7.13       | 7.13      | 7.03      | 7.23 | 0.00       |
| Aroclor-1260-5       | (5) | 7.37       | 7.37      | 7.27      | 7.47 | 0.00       |
| Tetrachloro-m-xylene |     | 3.77       | 3.77      | 3.67      | 3.87 | 0.00       |
| Decachlorobiphenyl   |     | 8.76       | 8.77      | 8.67      | 8.87 | 0.01       |