

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TULL02

Lab Code: ACE

SDG NO.: Q3790

Continuing Calib Date: 12/09/2025

Initial Calibration Date(s): 11/21/2025

11/21/2025

Continuing Calib Time: 04:52

Initial Calibration Time(s): 10:44

19:39

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.76	5.77	5.67	5.87	0.01
Aroclor-1242-2 (2)	5.79	5.79	5.69	5.89	0.00
Aroclor-1242-3 (3)	5.85	5.85	5.75	5.95	0.00
Aroclor-1242-4 (4)	5.95	5.95	5.85	6.05	0.00
Aroclor-1242-5 (5)	6.67	6.68	6.58	6.78	0.01
Tetrachloro-m-xylene	4.61	4.62	4.52	4.72	0.01
Decachlorobiphenyl	10.36	10.37	10.27	10.47	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TULL02

Lab Code: ACE

SDG NO.: Q3790

Continuing Calib Date: 12/09/2025

Initial Calibration Date(s): 11/21/2025

11/21/2025

Continuing Calib Time: 04:52

Initial Calibration Time(s): 10:44

19:39

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.85	4.85	4.75	4.95	0.00
Aroclor-1242-2	(2)	4.86	4.87	4.77	4.97	0.01
Aroclor-1242-3	(3)	5.04	5.05	4.95	5.15	0.01
Aroclor-1242-4	(4)	5.12	5.13	5.03	5.23	0.01
Aroclor-1242-5	(5)	5.64	5.65	5.55	5.75	0.01
Tetrachloro-m-xylene		3.77	3.77	3.67	3.87	0.00
Decachlorobiphenyl		8.76	8.77	8.67	8.87	0.01