

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TULL02

Lab Code: ACE

SDG NO.: Q3790

Continuing Calib Date: 12/09/2025

Initial Calibration Date(s): 11/21/2025

11/21/2025

Continuing Calib Time: 05:25

Initial Calibration Time(s): 10:44

19:39

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.61	6.62	6.52	6.72	0.01
Aroclor-1254-2	(2)	6.83	6.83	6.73	6.93	0.00
Aroclor-1254-3	(3)	7.19	7.20	7.10	7.30	0.01
Aroclor-1254-4	(4)	7.47	7.48	7.38	7.58	0.01
Aroclor-1254-5	(5)	7.89	7.89	7.79	7.99	0.00
Tetrachloro-m-xylene		4.61	4.62	4.52	4.72	0.01
Decachlorobiphenyl		10.36	10.37	10.27	10.47	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TULL02

Lab Code: ACE

SDG NO.: Q3790

Continuing Calib Date: 12/09/2025

Initial Calibration Date(s): 11/21/2025

11/21/2025

Continuing Calib Time: 05:25

Initial Calibration Time(s): 10:44

19:39

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.64	5.65	5.55	5.75	0.01
Aroclor-1254-2 (2)	5.79	5.80	5.70	5.90	0.01
Aroclor-1254-3 (3)	6.19	6.20	6.10	6.30	0.01
Aroclor-1254-4 (4)	6.42	6.43	6.33	6.53	0.01
Aroclor-1254-5 (5)	6.83	6.84	6.74	6.94	0.01
Tetrachloro-m-xylene	3.77	3.77	3.67	3.87	0.00
Decachlorobiphenyl	8.76	8.77	8.67	8.87	0.01