

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TULL02
Lab Code: ACE **SDG NO.:** Q3790
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 12/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076868.D **Time Analyzed:** 18:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.769	5.667	5.867	449.130	500.000	-10.2
Aroclor-1016-2	5.791	5.687	5.887	457.850	500.000	-8.4
Aroclor-1016-3	5.853	5.750	5.950	451.920	500.000	-9.6
Aroclor-1016-4	5.950	5.847	6.047	454.630	500.000	-9.1
Aroclor-1016-5	6.243	6.140	6.340	460.730	500.000	-7.9
Aroclor-1260-1	7.359	7.256	7.456	445.100	500.000	-11.0
Aroclor-1260-2	7.612	7.509	7.709	431.320	500.000	-13.7
Aroclor-1260-3	7.969	7.867	8.067	440.420	500.000	-11.9
Aroclor-1260-4	8.196	8.093	8.293	436.100	500.000	-12.8
Aroclor-1260-5	8.519	8.416	8.616	429.350	500.000	-14.1
Decachlorobiphenyl	10.366	10.265	10.465	43.760	50.000	-12.5
Tetrachloro-m-xylene	4.618	4.514	4.714	45.080	50.000	-9.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TULL02
Lab Code: ACE **SDG NO.:** Q3790
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 12/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076868.D **Time Analyzed:** 18:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.849	4.750	4.950	459.660	500.000	-8.1
Aroclor-1016-2	4.867	4.767	4.967	463.270	500.000	-7.3
Aroclor-1016-3	5.042	4.942	5.142	464.570	500.000	-7.1
Aroclor-1016-4	5.084	4.984	5.184	463.110	500.000	-7.4
Aroclor-1016-5	5.296	5.197	5.397	463.300	500.000	-7.3
Aroclor-1260-1	6.323	6.224	6.424	456.540	500.000	-8.7
Aroclor-1260-2	6.510	6.412	6.612	452.200	500.000	-9.6
Aroclor-1260-3	6.663	6.564	6.764	447.010	500.000	-10.6
Aroclor-1260-4	7.131	7.033	7.233	445.490	500.000	-10.9
Aroclor-1260-5	7.372	7.273	7.473	439.920	500.000	-12.0
Decachlorobiphenyl	8.762	8.666	8.866	41.610	50.000	-16.8
Tetrachloro-m-xylene	3.772	3.671	3.871	46.870	50.000	-6.3