

Surrogate Summary

SDG No.: Q3790

Client: Tully Construction Co., Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PP076598.D	PIBLK-PP076598.D	Tetrachloro-m-xyl	1	20	22.0	110		60	140
		Decachlorobiphen	1	20	23.3	117		60	140
		Tetrachloro-m-xyl	2	20	21.3	106		60	140
		Decachlorobiphen	2	20	22.1	110		60	140
I.BLK-PP076857.D	PIBLK-PP076857.D	Tetrachloro-m-xyl	1	20	18.9	94		60	140
		Decachlorobiphen	1	20	18.4	92		60	140
		Tetrachloro-m-xyl	2	20	17.2	86		60	140
		Decachlorobiphen	2	20	15.9	80		60	140
PB170862BL	PB170862BL	Tetrachloro-m-xyl	1	20	23.0	115		21	165
		Decachlorobiphen	1	20	22.3	111		10	170
		Tetrachloro-m-xyl	2	20	22.2	111		21	165
		Decachlorobiphen	2	20	20.2	101		10	170
PB170862BS	PB170862BS	Tetrachloro-m-xyl	1	20	23.8	119		21	165
		Decachlorobiphen	1	20	22.6	113		10	170
		Tetrachloro-m-xyl	2	20	21.4	107		21	165
		Decachlorobiphen	2	20	20.2	101		10	170
Q3790-01	304 FURMAN SOIL	Tetrachloro-m-xyl	1	20	21.4	107		21	165
		Decachlorobiphen	1	20	14.1	70		10	170
		Tetrachloro-m-xyl	2	20	19.1	96		21	165
		Decachlorobiphen	2	20	12.4	62		10	170
Q3790-01MS	304 FURMAN SOILMS	Tetrachloro-m-xyl	1	20	18.9	94		21	165
		Decachlorobiphen	1	20	14.2	71		10	170
		Tetrachloro-m-xyl	2	20	16.3	82		21	165
		Decachlorobiphen	2	20	12.4	62		10	170
Q3790-01MSD	304 FURMAN SOILMSD	Tetrachloro-m-xyl	1	20	20.2	101		21	165
		Decachlorobiphen	1	20	14.8	74		10	170
		Tetrachloro-m-xyl	2	20	17.7	88		21	165
		Decachlorobiphen	2	20	13.1	65		10	170
I.BLK-PP076872.D	PIBLK-PP076872.D	Tetrachloro-m-xyl	1	20	17.9	89		60	140
		Decachlorobiphen	1	20	17.9	90		60	140
		Tetrachloro-m-xyl	2	20	16.8	84		60	140
		Decachlorobiphen	2	20	16.0	80		60	140