

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Alliance

Contract: TULL02

ProjectID: MTA 26 Stations - Pierrepont

Lab Code: ACE

SDG No.: Q3790

Calibration Sequence : FF120125		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	166514422	97950	FF016829.D
850	87492081	102932	FF016830.D
340	36632833	107744	FF016831.D
170	18022029	106012	FF016832.D
85	9222027	108494	FF016833.D
AVG RF : 104626		% RSD : 4.112	AVG RT : 15.0426