

**TPH GC INITIAL CALIBRATION SUMMARY**

Lab Name: Alliance

Contract: TULL02

ProjectID: MTA 26 Stations - Pierrepont

Lab Code: ACE

SDG No.: Q3790

| Calibration Sequence : FG120125 |            | Test : TPH GC    |                  |
|---------------------------------|------------|------------------|------------------|
| Concentration (PPM)             | Area Count | Reference Factor | File ID          |
| 1700                            | 215295675  | 126645           | FG017031.D       |
| 850                             | 89629765   | 105447           | FG017032.D       |
| 340                             | 36738120   | 108053           | FG017033.D       |
| 170                             | 17526372   | 103096           | FG017034.D       |
| 85                              | 8474580    | 99701            | FG017035.D       |
| AVG RF : 108588                 |            | % RSD : 9.717    | AVG RT : 15.1388 |