

Analytical Sequence

Client: Tully Construction Co., Inc.

SDG No.: Q3790

Project: MTA 26 Stations - Pierrepont

Instrument ID: FID_G

GC Column: RXI-1MS **ID:** 0.18 (mm)

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

MEAN SUROGATE RT FROM INITIAL CALIBRATION 15.0426					
CLIENT SAMPLE NO.	LAB SAMPLE ID	DATE AND TIME ANALYZED	DATAFILE	RT	#
PIBLK01	LBLK01	10 Dec 2025 08:09	FF016884.D	15.041	
50 PPM TRPH STD	50 PPM TRPH STD	10 Dec 2025 09:38	FF016885.D	15.044	
PB170872BL	PB170872BL	10 Dec 2025 10:37	FF016887.D	15.042	
PB170872BS	PB170872BS	10 Dec 2025 11:08	FF016888.D	15.041	
304 FURMAN SOIL-TPH-1	Q3790-04	10 Dec 2025 11:37	FF016889.D	15.041	
304 FURMAN SOIL-TPH-1MS	Q3790-04MS	10 Dec 2025 12:07	FF016890.D	15.042	
304 FURMAN SOIL-TPH-1MSD	Q3790-04MSD	10 Dec 2025 12:36	FF016891.D	15.042	
PIBLK02	LBLK02	10 Dec 2025 13:06	FF016892.D	15.042	
50 PPM TRPH STD	50 PPM TRPH STD	10 Dec 2025 13:35	FF016893.D	15.045	
PIBLK03	LBLK03	10 Dec 2025 08:09	FG017150.D	15.136	
50 PPM TRPH STD	50 PPM TRPH STD	10 Dec 2025 08:38	FG017151.D	15.135	
304 FURMAN SOIL	Q3790-01	10 Dec 2025 09:38	FG017153.D	15.134	
304 FURMAN SOIL-TPH-2	Q3790-05	10 Dec 2025 10:07	FG017154.D	15.134	
304 FURMAN SOIL-TPH-3	Q3790-06	10 Dec 2025 10:37	FG017155.D	15.134	
PIBLK04	LBLK04	10 Dec 2025 11:08	FG017156.D	15.136	
50 PPM TRPH STD	50 PPM TRPH STD	10 Dec 2025 11:37	FG017157.D	15.137	