

SOIL TPH GC SURROGATE RECOVERY

Lab Name: Alliance

Client: Tully Construction Co., Inc.

Lab Code: ACE

SDG No.: Q3790

CLIENT ID	S1 TETRACOSANE-d50	S2	S3	S4	TOT OUT
PIBLK-FF016884.D	96				0
PIBLK-FF016892.D	92				0
PIBLK-FG017150.D	80				0
PIBLK-FG017156.D	78				0
PB170872BL	92				0
PB170872BS	95				0
304 FURMAN SOIL	49				0
304 FURMAN SOIL-TPH-1	79				0
304 FURMAN SOIL-TPH-1MS	59				0
304 FURMAN SOIL-TPH-1MSD	57				0
304 FURMAN SOIL-TPH-2	53				0
304 FURMAN SOIL-TPH-3	39				0

QC LIMITS

TETRACOSANE-d50

For Water : 29-130

For Soil : 37-130

Column to be used to flag recovery values

* Values outside of contract required QC limits

D Surrogate Diluted Out