

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120925\
 Data File : BP026258.D
 Acq On : 09 Dec 2025 12:12
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 09 12:31:40 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 05 14:52:44 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	83	0.00
2	1,4-Dioxane	40.000	41.821	-4.6	90	0.00
3	Pyridine	40.000	40.256	-0.6	87	0.00
4	n-Nitrosodimethylamine	40.000	39.361	1.6	86	0.00
5 S	2-Fluorophenol	80.000	80.061	-0.1	84	0.00
6	Aniline	40.000	37.196	7.0	84	0.00
7 S	Phenol-d6	80.000	76.631	4.2	83	0.00
8	2-Chlorophenol	40.000	39.642	0.9	83	0.00
9	Benzaldehyde	40.000	47.567	-18.9	82	0.00
10 C	Phenol	40.000	38.383	4.0	84	0.00
11	bis(2-Chloroethyl)ether	40.000	38.283	4.3	83	0.00
12	1,3-Dichlorobenzene	40.000	41.441	-3.6	86	0.00
13 C	1,4-Dichlorobenzene	40.000	40.870	-2.2	84	0.00
14	1,2-Dichlorobenzene	40.000	40.988	-2.5	85	0.00
15	Benzyl Alcohol	40.000	38.604	3.5	86	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.012	5.0	86	0.00
17	2-Methylphenol	40.000	38.699	3.3	85	0.00
18	Hexachloroethane	40.000	41.922	-4.8	87	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.428	3.9	87	0.00
20	3+4-Methylphenols	40.000	37.786	5.5	86	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	81	0.00
22	Acetophenone	40.000	41.337	-3.3	85	0.00
23 S	Nitrobenzene-d5	80.000	85.398	-6.7	87	0.00
24	Nitrobenzene	40.000	41.908	-4.8	86	0.00
25	Isophorone	40.000	41.742	-4.4	90	0.00
26 C	2-Nitrophenol	40.000	43.361	-8.4	85	0.00
27	2,4-Dimethylphenol	40.000	39.428	1.4	91	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.482	-1.2	86	0.00
29 C	2,4-Dichlorophenol	40.000	42.647	-6.6	86	0.00
30	1,2,4-Trichlorobenzene	40.000	43.440	-8.6	85	0.00
31	Naphthalene	40.000	41.663	-4.2	85	0.00
32	Benzoic acid	40.000	45.533	-13.8	145	0.04
33	4-Chloroaniline	40.000	41.012	-2.5	89	0.00
34 C	Hexachlorobutadiene	40.000	44.682	-11.7	85	0.00
35	Caprolactam	40.000	40.224	-0.6	88	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.392	-6.0	91	0.00
37	2-Methylnaphthalene	40.000	41.902	-4.8	87	0.00
38	1-Methylnaphthalene	40.000	42.331	-5.8	87	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	84	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	43.691	-9.2	88	0.00
41 P	Hexachlorocyclopentadiene	40.000	44.519	-11.3	117	0.00
42 S	2,4,6-Tribromophenol	80.000	90.210	-12.8	92	-0.01
43 C	2,4,6-Trichlorophenol	40.000	43.279	-8.2	90	0.00
44	2,4,5-Trichlorophenol	40.000	43.489	-8.7	92	-0.01
45 S	2-Fluorobiphenyl	80.000	85.590	-7.0	88	-0.01
46	1,1'-Biphenyl	40.000	41.800	-4.5	88	-0.01
47	2-Chloronaphthalene	40.000	41.899	-4.7	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120925\
 Data File : BP026258.D
 Acq On : 09 Dec 2025 12:12
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 09 12:31:40 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 05 14:52:44 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.828	-4.6	92	0.00
49	Acenaphthylene	40.000	42.077	-5.2	89	0.00
50	Dimethylphthalate	40.000	42.663	-6.7	91	-0.02
51	2,6-Dinitrotoluene	40.000	45.130	-12.8	92	-0.01
52 C	Acenaphthene	40.000	41.261	-3.2	89	-0.01
53	3-Nitroaniline	40.000	40.470	-1.2	88	-0.02
54 P	2,4-Dinitrophenol	40.000	54.736	-36.8#	324	0.00
55	Dibenzofuran	40.000	42.175	-5.4	89	0.00
56 P	4-Nitrophenol	40.000	39.466	1.3	95	0.00
57	2,4-Dinitrotoluene	40.000	44.292	-10.7	93	0.00
58	Fluorene	40.000	42.470	-6.2	90	-0.01
59	2,3,4,6-Tetrachlorophenol	40.000	44.192	-10.5	94	-0.02
60	Diethylphthalate	40.000	42.548	-6.4	92	0.00
61	4-Chlorophenyl-phenylether	40.000	43.517	-8.8	90	0.00
62	4-Nitroaniline	40.000	40.217	-0.5	91	-0.01
63	Azobenzene	40.000	41.651	-4.1	91	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	84	0.00
65	4,6-Dinitro-2-methylphenol	40.000	49.538	-23.8	153	-0.01
66 c	n-Nitrosodiphenylamine	40.000	42.940	-7.3	90	-0.01
67	4-Bromophenyl-phenylether	40.000	45.107	-12.8	91	-0.01
68	Hexachlorobenzene	40.000	45.735	-14.3	92	0.00
69	Atrazine	40.000	43.086	-7.7	91	-0.01
70 C	Pentachlorophenol	40.000	44.133	-10.3	94	-0.01
71	Phenanthrene	40.000	42.598	-6.5	89	0.00
72	Anthracene	40.000	43.406	-8.5	90	0.00
73	Carbazole	40.000	41.391	-3.5	88	-0.01
74	Di-n-butylphthalate	40.000	43.084	-7.7	90	-0.02
75 C	Fluoranthene	40.000	42.345	-5.9	88	-0.01
76 I	Chrysene-d12	20.000	20.000	0.0	80	-0.01
77	Benzydine	40.000	36.135	9.7	79	0.00
78	Pyrene	40.000	42.284	-5.7	86	-0.01
79 S	Terphenyl-d14	80.000	86.966	-8.7	87	-0.01
80	Butylbenzylphthalate	40.000	42.230	-5.6	86	0.00
81	Benzo(a)anthracene	40.000	40.785	-2.0	82	-0.01
82	3,3'-Dichlorobenzidine	40.000	40.348	-0.9	81	-0.01
83	Chrysene	40.000	42.015	-5.0	83	-0.01
84	Bis(2-ethylhexyl)phthalate	40.000	41.849	-4.6	86	0.00
85 c	Di-n-octyl phthalate	40.000	41.267	-3.2	83	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	76	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	41.785	-4.5	75	-0.05
88	Benzo(b)fluoranthene	40.000	41.769	-4.4	79	-0.02
89	Benzo(k)fluoranthene	40.000	41.034	-2.6	78	-0.02
90 C	Benzo(a)pyrene	40.000	41.478	-3.7	78	-0.03
91	Dibenzo(a,h)anthracene	40.000	41.618	-4.0	75	-0.02
92	Benzo(g,h,i)perylene	40.000	41.906	-4.8	75	-0.06

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120925\
Data File : BP026258.D
Acq On : 09 Dec 2025 12:12
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Dec 09 12:31:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 05 14:52:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0