

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120925\
 Data File : BP026258.D
 Acq On : 09 Dec 2025 12:12
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 09 12:31:40 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 05 14:52:44 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
2	1,4-Dioxane	0.506	0.529	-4.5	90	0.00
3	Pyridine	1.453	1.462	-0.6	87	0.00
4	n-Nitrosodimethylamine	0.542	0.533	1.7	86	0.00
5 S	2-Fluorophenol	1.244	1.245	-0.1	84	0.00
6	Aniline	2.061	1.916	7.0	84	0.00
7 S	Phenol-d6	1.570	1.504	4.2	83	0.00
8	2-Chlorophenol	1.414	1.401	0.9	83	0.00
9	Benzaldehyde	0.847	1.007	-18.9	82	0.00
10 C	Phenol	1.689	1.621	4.0	84	0.00
11	bis(2-Chloroethyl)ether	1.324	1.267	4.3	83	0.00
12	1,3-Dichlorobenzene	1.530	1.585	-3.6	86	0.00
13 C	1,4-Dichlorobenzene	1.541	1.575	-2.2	84	0.00
14	1,2-Dichlorobenzene	1.480	1.516	-2.4	85	0.00
15	Benzyl Alcohol	1.142	1.102	3.5	86	0.00
16	2,2'-oxybis(1-Chloropropane	1.848	1.757	4.9	86	0.00
17	2-Methylphenol	1.146	1.109	3.2	85	0.00
18	Hexachloroethane	0.560	0.587	-4.8	87	0.00
19 P	n-Nitroso-di-n-propylamine	0.968	0.930	3.9	87	0.00
20	3+4-Methylphenols	1.583	1.495	5.6	86	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	81	0.00
22	Acetophenone	0.507	0.524	-3.4	85	0.00
23 S	Nitrobenzene-d5	0.351	0.375	-6.8	87	0.00
24	Nitrobenzene	0.355	0.372	-4.8	86	0.00
25	Isophorone	0.689	0.719	-4.4	90	0.00
26 C	2-Nitrophenol	0.180	0.195	-8.3	85	0.00
27	2,4-Dimethylphenol	0.318	0.313	1.6	91	0.00
28	bis(2-Chloroethoxy)methane	0.433	0.438	-1.2	86	0.00
29 C	2,4-Dichlorophenol	0.324	0.345	-6.5	86	0.00
30	1,2,4-Trichlorobenzene	0.333	0.362	-8.7	85	0.00
31	Naphthalene	1.081	1.126	-4.2	85	0.00
32	Benzoic acid	0.255	0.290	-13.7	145	0.04
33	4-Chloroaniline	0.453	0.464	-2.4	89	0.00
34 C	Hexachlorobutadiene	0.217	0.243	-12.0	85	0.00
35	Caprolactam	0.114	0.115	-0.9	88	0.00
36 C	4-Chloro-3-methylphenol	0.341	0.361	-5.9	91	0.00
37	2-Methylnaphthalene	0.762	0.798	-4.7	87	0.00
38	1-Methylnaphthalene	0.744	0.788	-5.9	87	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	84	0.00
40	1,2,4,5-Tetrachlorobenzene	0.611	0.667	-9.2	88	0.00
41 P	Hexachlorocyclopentadiene	0.381	0.424	-11.3	117	0.00
42 S	2,4,6-Tribromophenol	0.260	0.293	-12.7	92	-0.01
43 C	2,4,6-Trichlorophenol	0.415	0.449	-8.2	90	0.00
44	2,4,5-Trichlorophenol	0.462	0.502	-8.7	92	-0.01
45 S	2-Fluorobiphenyl	1.376	1.472	-7.0	88	-0.01
46	1,1'-Biphenyl	1.510	1.578	-4.5	88	-0.01
47	2-Chloronaphthalene	1.190	1.246	-4.7	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120925\
 Data File : BP026258.D
 Acq On : 09 Dec 2025 12:12
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 09 12:31:40 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 05 14:52:44 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.326	0.341	-4.6	92	0.00
49	Acenaphthylene	1.953	2.054	-5.2	89	0.00
50	Dimethylphthalate	1.490	1.589	-6.6	91	-0.02
51	2,6-Dinitrotoluene	0.295	0.332	-12.5	92	-0.01
52 C	Acenaphthene	1.142	1.178	-3.2	89	-0.01
53	3-Nitroaniline	0.344	0.348	-1.2	88	-0.02
54 P	2,4-Dinitrophenol	0.157	0.215	-36.9#	324#	0.00
55	Dibenzofuran	1.776	1.872	-5.4	89	0.00
56 P	4-Nitrophenol	0.307	0.303	1.3	95	0.00
57	2,4-Dinitrotoluene	0.438	0.485	-10.7	93	0.00
58	Fluorene	1.405	1.491	-6.1	90	-0.01
59	2,3,4,6-Tetrachlorophenol	0.391	0.432	-10.5	94	-0.02
60	Diethylphthalate	1.495	1.590	-6.4	92	0.00
61	4-Chlorophenyl-phenylether	0.700	0.762	-8.9	90	0.00
62	4-Nitroaniline	0.357	0.359	-0.6	91	-0.01
63	Azobenzene	1.257	1.309	-4.1	91	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	84	0.00
65	4,6-Dinitro-2-methylphenol	0.118	0.146	-23.7	153#	-0.01
66 c	n-Nitrosodiphenylamine	0.618	0.664	-7.4	90	-0.01
67	4-Bromophenyl-phenylether	0.222	0.250	-12.6	91	-0.01
68	Hexachlorobenzene	0.258	0.295	-14.3	92	0.00
69	Atrazine	0.233	0.250	-7.3	91	-0.01
70 C	Pentachlorophenol	0.179	0.197	-10.1	94	-0.01
71	Phenanthrene	1.127	1.201	-6.6	89	0.00
72	Anthracene	1.148	1.246	-8.5	90	0.00
73	Carbazole	1.087	1.125	-3.5	88	-0.01
74	Di-n-butylphthalate	1.284	1.383	-7.7	90	-0.02
75 C	Fluoranthene	1.372	1.453	-5.9	88	-0.01
76 I	Chrysene-d12	1.000	1.000	0.0	80	-0.01
77	Benzydine	0.628	0.567	9.7	79	0.00
78	Pyrene	1.303	1.378	-5.8	86	-0.01
79 S	Terphenyl-d14	0.979	1.064	-8.7	87	-0.01
80	Butylbenzylphthalate	0.572	0.604	-5.6	86	0.00
81	Benzo(a)anthracene	1.372	1.399	-2.0	82	-0.01
82	3,3'-Dichlorobenzidine	0.499	0.504	-1.0	81	-0.01
83	Chrysene	1.299	1.364	-5.0	83	-0.01
84	Bis(2-ethylhexyl)phthalate	0.860	0.900	-4.7	86	0.00
85 c	Di-n-octyl phthalate	1.517	1.565	-3.2	83	-0.02
86 I	Perylene-d12	1.000	1.000	0.0	76	-0.01
87	Indeno(1,2,3-cd)pyrene	1.512	1.579	-4.4	75	-0.05
88	Benzo(b)fluoranthene	1.215	1.269	-4.4	79	-0.02
89	Benzo(k)fluoranthene	1.216	1.247	-2.5	78	-0.02
90 C	Benzo(a)pyrene	1.153	1.195	-3.6	78	-0.03
91	Dibenzo(a,h)anthracene	1.236	1.286	-4.0	75	-0.02
92	Benzo(g,h,i)perylene	1.232	1.291	-4.8	75	-0.06

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120925\
Data File : BP026258.D
Acq On : 09 Dec 2025 12:12
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Dec 09 12:31:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 05 14:52:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0