

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3790

Analytical Method: SW8270E

Client: Tully Construction Co., Inc.

DataFile: BP026262.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q3795-01MS		Client Sample ID: 72-11966MS									
Naphthalene	1100	0	1000	ug/Kg	91				66	113	
Acenaphthylene	1100	0	1100	ug/Kg	100				69	128	
Acenaphthene	1100	0	1100	ug/Kg	100				70	121	
Fluorene	1100	0	1100	ug/Kg	100				67	114	
Phenanthrene	1100	0	1100	ug/Kg	100				52	128	
Anthracene	1100	0	1100	ug/Kg	100				62	124	
Fluoranthene	1100	0	1100	ug/Kg	100				31	152	
Pyrene	1100	0	1100	ug/Kg	100				37	122	
Benzo(a)anthracene	1100	0	1100	ug/Kg	100				53	119	
Chrysene	1100	0	1100	ug/Kg	100				57	121	
Benzo(b)fluoranthene	1100	0	1100	ug/Kg	100				52	117	
Benzo(k)fluoranthene	1100	0	1100	ug/Kg	100				57	134	
Benzo(a)pyrene	1100	0	1100	ug/Kg	100				70	142	
Indeno(1,2,3-cd)pyrene	1100	0	1100	ug/Kg	100				48	129	
Dibenz(a,h)anthracene	1100	0	1100	ug/Kg	100				43	123	
Benzo(g,h,i)perylene	1100	0	1100	ug/Kg	100				40	133	