

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3790

Client ID : SSTDCCC040

Date Analyzed: 12/09/2025

Lab File ID: BG064902.D

Time Analyzed: 12:49

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	22681	8.155	86644	10.97	75628	14.76
UPPER LIMIT	45362	8.655	173288	11.469	151256	15.264
LOWER LIMIT	11340.5	7.655	43322	10.469	37814	14.264
EPA SAMPLE NO.						
01 304 FURMAN SOIL	22406	8.15	90673	10.97	79761	14.76

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.