

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3790

Client ID : SSTDCCC040

Date Analyzed: 12/09/2025

Lab File ID: BP026258.D

Time Analyzed: 12:12

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	228234	7.655	860796	10.42	547767	14.28
UPPER LIMIT	456468	8.155	1721590	10.919	1095530	14.784
LOWER LIMIT	114117	7.155	430398	9.919	273884	13.784
EPA SAMPLE NO.						
01 PB170865BL	140237	7.66	530456	10.41	323504	14.28
02 PB170865BS	233832	7.65	945889	10.41	611619	14.28
03 72-11966MS	209963	7.65	834686	10.42	541018	14.28
04 72-11966MSD	173541	7.66	672715	10.41	419939	14.28

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.