

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BG064901.D
Instrument ID: BNA_G

Contract: TULL02
SDG NO.: Q3790
DFTPP Injection Date: 12/09/2025
DFTPP Injection Time: 11:26

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.3 (1.5) 1
69	Mass 69 relative abundance	23.1
70	Less than 2.0% of mass 69	0.0 (0.0) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.4
365	Greater than 1% of mass 198	6.3
441	Present, but less than mass 443	15.7
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.4 (18.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BG064902.D	12/09/2025	12:49
304 FURMAN SOIL	Q3790-01	BG064914.D	12/09/2025	21:05