

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3790</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Tully Construction Co., Inc.</u>	Datafile :	<u>VW032589.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW1208SBS01	Dichlorodifluoromethane	20	22.0	ug/Kg	110	3		64	136	20
	Chloromethane	20	18.9	ug/Kg	95	1		73	129	15
	Vinyl chloride	20	20.9	ug/Kg	104	4		73	133	15
	Bromomethane	20	19.9	ug/Kg	100	2		77	137	15
	Chloroethane	20	19.8	ug/Kg	99	1		69	130	15
	Trichlorofluoromethane	20	21.1	ug/Kg	106	2		69	134	27
	1,1,2-Trichlorotrifluoroethane	20	21.9	ug/Kg	110	7		81	123	20
	1,1-Dichloroethene	20	20.0	ug/Kg	100	2		79	121	16
	Acetone	100	110	ug/Kg	110	0		60	174	28
	Carbon disulfide	20	19.6	ug/Kg	98	2		73	125	15
	Methyl tert-butyl Ether	20	20.7	ug/Kg	104	0		77	129	20
	Methyl Acetate	20	20.1	ug/Kg	101	3		51	140	30
	Methylene Chloride	20	20.6	ug/Kg	103	6		54	158	20
	trans-1,2-Dichloroethene	20	19.8	ug/Kg	99	1		80	123	15
	1,1-Dichloroethane	20	19.8	ug/Kg	99	2		82	123	15
	Cyclohexane	20	20.6	ug/Kg	103	2		76	122	15
	2-Butanone	100	110	ug/Kg	110	10		69	131	29
	Carbon Tetrachloride	20	19.8	ug/Kg	99	1		76	129	15
	cis-1,2-Dichloroethene	20	19.4	ug/Kg	97	1		82	123	15
	Bromochloromethane	20	29.0	ug/Kg	145	34	* *	80	127	15
	Chloroform	20	20.0	ug/Kg	100	0		82	125	15
	1,1,1-Trichloroethane	20	20.1	ug/Kg	101	1		80	126	15
	Methylcyclohexane	20	20.5	ug/Kg	103	6		77	123	15
	Benzene	20	19.6	ug/Kg	98	2		84	121	15
	1,2-Dichloroethane	20	20.5	ug/Kg	103	3		81	126	15
	Trichloroethene	20	19.4	ug/Kg	97	0		83	122	15
	1,2-Dichloropropane	20	20.2	ug/Kg	101	1		83	122	15
	Bromodichloromethane	20	19.8	ug/Kg	99	0		82	123	15
	4-Methyl-2-Pentanone	100	110	ug/Kg	110	0		70	135	26
	Toluene	20	20.4	ug/Kg	102	3		83	122	15
	t-1,3-Dichloropropene	20	19.8	ug/Kg	99	3		78	124	15
	cis-1,3-Dichloropropene	20	20.1	ug/Kg	101	4		81	122	15
	1,1,2-Trichloroethane	20	20.3	ug/Kg	102	1		82	125	15
	2-Hexanone	100	110	ug/Kg	110	0		66	138	27
	Dibromochloromethane	20	20.2	ug/Kg	101	4		79	125	15
	1,2-Dibromoethane	20	20.4	ug/Kg	102	1		80	125	16
	Tetrachloroethene	20	19.9	ug/Kg	100	9		83	125	15
	Chlorobenzene	20	20.3	ug/Kg	102	13		84	122	15
	Ethyl Benzene	20	20.4	ug/Kg	102	11		82	124	15
	m/p-Xylenes	40	41.4	ug/Kg	104	10		83	124	15
	o-Xylene	20	20.7	ug/Kg	104	14		83	123	15
	Styrene	20	20.9	ug/Kg	104	7		82	124	15
	Bromoform	20	20.8	ug/Kg	104	8		75	127	16
	Isopropylbenzene	20	20.4	ug/Kg	102	8		82	124	15
	1,1,2,2-Tetrachloroethane	20	21.8	ug/Kg	109	6		77	127	21
	1,3-Dichlorobenzene	20	20.3	ug/Kg	102	5		83	122	15
	1,4-Dichlorobenzene	20	19.8	ug/Kg	99	4		84	121	15
	1,2-Dichlorobenzene	20	21.1	ug/Kg	106	7		83	124	15

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3790</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Tully Construction Co., Inc.</u>	Datafile :	<u>VW032589.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW1208SBSD01	1,2-Dibromo-3-Chloropropane	20	20.7	ug/Kg	104	5		66	134	20
	1,2,4-Trichlorobenzene	20	19.6	ug/Kg	98	13		78	127	20
	1,2,3-Trichlorobenzene	20	19.5	ug/Kg	98	9		70	137	16