

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TULL02
 Lab Code: ACE SDG NO.: Q3790
 Lab File ID: VW032583.D Date Analyzed: 12/08/2025
 Instrument ID: MSVOA_W Time Analyzed: 09:40
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	354482	7.96	643883	8.85	585250	11.64
UPPER LIMIT	708964	8.459	1287770	9.349	1170500	12.135
LOWER LIMIT	177241	7.459	321942	8.349	292625	11.135
EPA SAMPLE NO.						
304 FURMAN SOIL	148034 *	7.96	272134 *	8.85	235198 *	11.64
VW1208SBL01	314569	7.97	616405	8.85	548609	11.63
VW1208SBS01	346269	7.97	656064	8.86	619620	11.64
VW1208SBSD01	314458	7.97	582852	8.85	505370	11.63

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.