

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TULL02
 Lab Code: ACE SDG NO.: Q3790
 Lab File ID: VW032583.D Date Analyzed: 12/08/2025
 Instrument ID: MSVOA_W Time Analyzed: 09:40
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	270635	13.55				
UPPER LIMIT	541270	14.05				
LOWER LIMIT	135318	13.05				
EPA SAMPLE NO.						
304 FURMAN SOIL	77597 *	13.56				
VW1208SBL01	222458	13.56				
VW1208SBS01	286140	13.55				
VW1208SBSD01	246600	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.