

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: TULL02

Lab Code: ACE

SDG NO.: Q3790

Lab File ID: VW032582.D

BFB Injection Date: 12/08/2025

Instrument ID: MSVOA_W

BFB Injection Time: 08:11

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.1
75	30.0 - 60.0% of mass 95	51.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.0 (0.0) 1
174	50.0 - 100.0% of mass 95	67.5
175	5.0 - 9.0% of mass 174	5.4 (7.9) 1
176	95.0 - 101.0% of mass 174	64.1 (95.1) 1
177	5.0 - 9.0% of mass 176	4.4 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VW032583.D	12/08/2025	09:40
VW1208SBL01	VW1208SBL01	VW032584.D	12/08/2025	10:18
VW1208SBS01	VW1208SBS01	VW032585.D	12/08/2025	11:01
VW1208SBSD01	VW1208SBSD01	VW032589.D	12/08/2025	12:39
304 FURMAN SOIL	Q3790-03	VW032592.D	12/08/2025	13:44