

Initial and Continuing Calibration Verification

Client: Europastry

SDG No.: Q3793

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

RunNo.: LB138149

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.01	7	100	90-110	12/08/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	12/08/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	12/08/2025

Initial and Continuing Calibration Verification

Client: Europastry

SDG No.: Q3793

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

RunNo.: LB138209

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	48.381	50	97	95-105	12/12/2025
Sample ID: CCV1 COD	mg/L	48.381	50	97	95-105	12/12/2025
Sample ID: CCV2 COD	mg/L	49.337	50	99	95-105	12/12/2025