

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3793
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB138133

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138133BS							
BOD5	mg/L	198	190		96	1	84.6-115.4	12/05/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3793
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB138150

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138150BS							
Oil and Grease	mg/L	20.0	16.7		84	1	78-114	12/08/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3793
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB138151

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138151BS							
TPH	mg/L	20.0	18.0		90	1	78-114	12/08/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3793
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB138151

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138151BSD							
TPH	mg/L	20.0	18.3		92	1	78-114	12/08/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3793
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB138190

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138190BS							
TSS	mg/L	550	577		105	1	90-110	12/11/2025

Laboratory Control Sample Summary

Client: Europastry

SDG No.: Q3793

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

Run No.: LB138209

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138209BS							
COD	mg/L	50	50.3		101	1	90-110	12/12/2025