



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Dal Tile - Sunnyvale Plant

SDG No.: Q3796

Contract: DALT03

Lab Code: ACE

Initial Calibration Source:

Continuing Calibration Source:

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	104	70 - 130	CV	12/08/2025	12:42	LB138142
CRI01	Aluminum	113	100	113	65 - 135	P	12/11/2025	11:50	LB138198
	Antimony	51.9	50.0	104	65 - 135	P	12/11/2025	11:50	LB138198
	Arsenic	21.2	20.0	106	65 - 135	P	12/11/2025	11:50	LB138198
	Barium	97.8	100	98	65 - 135	P	12/11/2025	11:50	LB138198
	Beryllium	6.39	6.0	106	65 - 135	P	12/11/2025	11:50	LB138198
	Boron	102	100	102	65 - 135	P	12/11/2025	11:50	LB138198
	Cadmium	5.86	6.0	98	65 - 135	P	12/11/2025	11:50	LB138198
	Calcium	2040	2000	102	65 - 135	P	12/11/2025	11:50	LB138198
	Chromium	10.3	10.0	103	65 - 135	P	12/11/2025	11:50	LB138198
	Cobalt	29.9	30.0	100	65 - 135	P	12/11/2025	11:50	LB138198
	Copper	21.8	20.0	109	65 - 135	P	12/11/2025	11:50	LB138198
	Iron	109	100	109	65 - 135	P	12/11/2025	11:50	LB138198
	Lead	13.7	12.0	114	65 - 135	P	12/11/2025	11:50	LB138198
	Magnesium	2130	2000	106	65 - 135	P	12/11/2025	11:50	LB138198
	Manganese	20.3	20.0	102	65 - 135	P	12/11/2025	11:50	LB138198
	Molybdenum	200	200	100	65 - 135	P	12/11/2025	11:50	LB138198
	Nickel	41.1	40.0	103	65 - 135	P	12/11/2025	11:50	LB138198
	Potassium	2030	2000	102	65 - 135	P	12/11/2025	11:50	LB138198
	Selenium	20.9	20.0	105	65 - 135	P	12/11/2025	11:50	LB138198
	Silver	11.3	10.0	113	65 - 135	P	12/11/2025	11:50	LB138198
	Sodium	1920	2000	96	65 - 135	P	12/11/2025	11:50	LB138198
	Thallium	39.1	40.0	98	65 - 135	P	12/11/2025	11:50	LB138198
	Tin	39.2	40.0	98	65 - 135	P	12/11/2025	11:50	LB138198
	Titanium	39.3	40.0	98	65 - 135	P	12/11/2025	11:50	LB138198
	Vanadium	41.2	40.0	103	65 - 135	P	12/11/2025	11:50	LB138198
	Zinc	43.9	40.0	110	65 - 135	P	12/11/2025	11:50	LB138198