



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Metals

- 3a -

#### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Dal Tile - Sunnyvale Plant

**SDG No.:** Q3796

**Contract:** DALT03

**Lab Code:** ACE

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|----|------------------|------------------|---------------|
| ICB142    | Mercury | 0.076          | +/-0.2              | U            | 0.20 | CV | 12/08/2025       | 12:35            | LB138142      |

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Dal Tile - Sunnyvale Plant

**SDG No.:** Q3796

**Contract:** DALT03

**Lab Code:** ACE

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|----|------------------|------------------|---------------|
| CCB74     | Mercury | 0.076          | +/-0.2              | U            | 0.20 | CV | 12/08/2025       | 12:40            | LB138142      |
| CCB75     | Mercury | 0.076          | +/-0.2              | U            | 0.20 | CV | 12/08/2025       | 13:10            | LB138142      |
| CCB76     | Mercury | 0.076          | +/-0.2              | U            | 0.20 | CV | 12/08/2025       | 13:37            | LB138142      |
| CCB77     | Mercury | 0.076          | +/-0.2              | U            | 0.20 | CV | 12/08/2025       | 13:58            | LB138142      |

## Metals

- 3a -

### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Dal Tile - Sunnyvale Plant

**SDG No.:** Q3796

**Contract:** DALT03

**Lab Code:** ACE

| Sample ID | Analyte    | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|------------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| ICB01     | Aluminum   | 11.3           | +/-50               | U            | 100  | P | 12/11/2025       | 11:46            | LB138198      |
|           | Antimony   | 6.76           | +/-25               | U            | 50.0 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Arsenic    | 5.12           | +/-10               | U            | 20.0 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Barium     | 14.6           | +/-50               | U            | 100  | P | 12/11/2025       | 11:46            | LB138198      |
|           | Beryllium  | 0.56           | +/-3                | U            | 6.00 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Boron      | 15.7           | +/-50               | U            | 100  | P | 12/11/2025       | 11:46            | LB138198      |
|           | Cadmium    | 0.50           | +/-3                | U            | 6.00 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Calcium    | 234            | +/-1000             | U            | 2000 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Chromium   | 2.12           | +/-5                | U            | 10.0 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Cobalt     | 2.26           | +/-15               | U            | 30.0 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Copper     | 4.60           | +/-10               | U            | 20.0 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Iron       | 23.4           | +/-50               | U            | 100  | P | 12/11/2025       | 11:46            | LB138198      |
|           | Lead       | 2.30           | +/-6                | U            | 12.0 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Magnesium  | 244            | +/-1000             | U            | 2000 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Manganese  | 5.94           | +/-10               | U            | 20.0 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Molybdenum | 18.7           | +/-100              | U            | 200  | P | 12/11/2025       | 11:46            | LB138198      |
|           | Nickel     | 3.06           | +/-20               | U            | 40.0 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Potassium  | 918            | +/-1000             | U            | 2000 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Selenium   | 9.64           | +/-10               | U            | 20.0 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Silver     | 1.62           | +/-5                | U            | 10.0 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Sodium     | 868            | +/-1000             | U            | 2000 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Thallium   | 4.38           | +/-20               | U            | 40.0 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Tin        | 5.14           | +/-20               | U            | 40.0 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Titanium   | 6.42           | +/-20               | U            | 40.0 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Vanadium   | 6.26           | +/-20               | U            | 40.0 | P | 12/11/2025       | 11:46            | LB138198      |
|           | Zinc       | 16.7           | +/-20               | U            | 40.0 | P | 12/11/2025       | 11:46            | LB138198      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Dal Tile - Sunnyvale Plant

**SDG No.:** Q3796

**Contract:** DALT03

**Lab Code:** ACE

| Sample ID | Analyte    | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|------------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB01     | Aluminum   | 11.3           | +/-50               | U            | 100  | P | 12/11/2025       | 12:18            | LB138198      |
|           | Antimony   | 6.76           | +/-25               | U            | 50.0 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Arsenic    | 5.12           | +/-10               | U            | 20.0 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Barium     | 14.6           | +/-50               | U            | 100  | P | 12/11/2025       | 12:18            | LB138198      |
|           | Beryllium  | 0.56           | +/-3                | U            | 6.00 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Boron      | 15.7           | +/-50               | U            | 100  | P | 12/11/2025       | 12:18            | LB138198      |
|           | Cadmium    | 0.50           | +/-3                | U            | 6.00 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Calcium    | 234            | +/-1000             | U            | 2000 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Chromium   | 2.12           | +/-5                | U            | 10.0 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Cobalt     | 2.26           | +/-15               | U            | 30.0 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Copper     | 4.60           | +/-10               | U            | 20.0 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Iron       | 23.4           | +/-50               | U            | 100  | P | 12/11/2025       | 12:18            | LB138198      |
|           | Lead       | 2.30           | +/-6                | U            | 12.0 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Magnesium  | 244            | +/-1000             | U            | 2000 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Manganese  | 5.94           | +/-10               | U            | 20.0 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Molybdenum | 18.7           | +/-100              | U            | 200  | P | 12/11/2025       | 12:18            | LB138198      |
|           | Nickel     | 3.06           | +/-20               | U            | 40.0 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Potassium  | 918            | +/-1000             | U            | 2000 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Selenium   | 9.64           | +/-10               | U            | 20.0 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Silver     | 1.62           | +/-5                | U            | 10.0 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Sodium     | 868            | +/-1000             | U            | 2000 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Thallium   | 4.38           | +/-20               | U            | 40.0 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Tin        | 5.14           | +/-20               | U            | 40.0 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Titanium   | 6.42           | +/-20               | U            | 40.0 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Vanadium   | 6.26           | +/-20               | U            | 40.0 | P | 12/11/2025       | 12:18            | LB138198      |
|           | Zinc       | 16.7           | +/-20               | U            | 40.0 | P | 12/11/2025       | 12:18            | LB138198      |
| CCB02     | Aluminum   | 11.3           | +/-50               | U            | 100  | P | 12/11/2025       | 13:08            | LB138198      |
|           | Antimony   | 6.76           | +/-25               | U            | 50.0 | P | 12/11/2025       | 13:08            | LB138198      |
|           | Arsenic    | 5.12           | +/-10               | U            | 20.0 | P | 12/11/2025       | 13:08            | LB138198      |
|           | Barium     | 14.6           | +/-50               | U            | 100  | P | 12/11/2025       | 13:08            | LB138198      |
|           | Beryllium  | 0.56           | +/-3                | U            | 6.00 | P | 12/11/2025       | 13:08            | LB138198      |
|           | Boron      | 15.7           | +/-50               | U            | 100  | P | 12/11/2025       | 13:08            | LB138198      |
|           | Cadmium    | 0.50           | +/-3                | U            | 6.00 | P | 12/11/2025       | 13:08            | LB138198      |
|           | Calcium    | 234            | +/-1000             | U            | 2000 | P | 12/11/2025       | 13:08            | LB138198      |
|           | Chromium   | 2.12           | +/-5                | U            | 10.0 | P | 12/11/2025       | 13:08            | LB138198      |
|           | Cobalt     | 2.26           | +/-15               | U            | 30.0 | P | 12/11/2025       | 13:08            | LB138198      |
|           | Copper     | 4.60           | +/-10               | U            | 20.0 | P | 12/11/2025       | 13:08            | LB138198      |
|           | Iron       | 23.4           | +/-50               | U            | 100  | P | 12/11/2025       | 13:08            | LB138198      |
|           | Lead       | 2.30           | +/-6                | U            | 12.0 | P | 12/11/2025       | 13:08            | LB138198      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Dal Tile - Sunnyvale Plant

**SDG No.:** Q3796

**Contract:** DALT03

**Lab Code:** ACE

| Sample ID | Analyte    | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|------------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB02     | Magnesium  | 244            | +/-1000             | U            | 2000 | P | 12/11/2025       | 13:08            | LB138198      |
|           | Manganese  | 5.94           | +/-10               | U            | 20.0 | P | 12/11/2025       | 13:08            | LB138198      |
|           | Molybdenum | 18.7           | +/-100              | U            | 200  | P | 12/11/2025       | 13:08            | LB138198      |
|           | Nickel     | 3.06           | +/-20               | U            | 40.0 | P | 12/11/2025       | 13:08            | LB138198      |
|           | Potassium  | 918            | +/-1000             | U            | 2000 | P | 12/11/2025       | 13:08            | LB138198      |
|           | Selenium   | 9.64           | +/-10               | U            | 20.0 | P | 12/11/2025       | 13:08            | LB138198      |
|           | Silver     | 1.62           | +/-5                | U            | 10.0 | P | 12/11/2025       | 13:08            | LB138198      |
|           | Sodium     | 868            | +/-1000             | U            | 2000 | P | 12/11/2025       | 13:08            | LB138198      |
|           | Thallium   | 4.38           | +/-20               | U            | 40.0 | P | 12/11/2025       | 13:08            | LB138198      |
|           | Tin        | 5.14           | +/-20               | U            | 40.0 | P | 12/11/2025       | 13:08            | LB138198      |
|           | Titanium   | 6.42           | +/-20               | U            | 40.0 | P | 12/11/2025       | 13:08            | LB138198      |
|           | Vanadium   | 6.26           | +/-20               | U            | 40.0 | P | 12/11/2025       | 13:08            | LB138198      |
| CCB03     | Zinc       | 16.7           | +/-20               | U            | 40.0 | P | 12/11/2025       | 13:08            | LB138198      |
|           | Aluminum   | 11.3           | +/-50               | U            | 100  | P | 12/11/2025       | 13:59            | LB138198      |
|           | Antimony   | 6.76           | +/-25               | U            | 50.0 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Arsenic    | 5.12           | +/-10               | U            | 20.0 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Barium     | 14.6           | +/-50               | U            | 100  | P | 12/11/2025       | 13:59            | LB138198      |
|           | Beryllium  | 0.56           | +/-3                | U            | 6.00 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Boron      | 15.7           | +/-50               | U            | 100  | P | 12/11/2025       | 13:59            | LB138198      |
|           | Cadmium    | 0.50           | +/-3                | U            | 6.00 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Calcium    | 234            | +/-1000             | U            | 2000 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Chromium   | 2.12           | +/-5                | U            | 10.0 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Cobalt     | 2.26           | +/-15               | U            | 30.0 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Copper     | 4.60           | +/-10               | U            | 20.0 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Iron       | 23.4           | +/-50               | U            | 100  | P | 12/11/2025       | 13:59            | LB138198      |
|           | Lead       | 2.30           | +/-6                | U            | 12.0 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Magnesium  | 244            | +/-1000             | U            | 2000 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Manganese  | 5.94           | +/-10               | U            | 20.0 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Molybdenum | 18.7           | +/-100              | U            | 200  | P | 12/11/2025       | 13:59            | LB138198      |
|           | Nickel     | 3.06           | +/-20               | U            | 40.0 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Potassium  | 918            | +/-1000             | U            | 2000 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Selenium   | 9.64           | +/-10               | U            | 20.0 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Silver     | 1.62           | +/-5                | U            | 10.0 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Sodium     | 868            | +/-1000             | U            | 2000 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Thallium   | 4.38           | +/-20               | U            | 40.0 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Tin        | 5.14           | +/-20               | U            | 40.0 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Titanium   | 6.42           | +/-20               | U            | 40.0 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Vanadium   | 6.26           | +/-20               | U            | 40.0 | P | 12/11/2025       | 13:59            | LB138198      |
|           | Zinc       | 16.7           | +/-20               | U            | 40.0 | P | 12/11/2025       | 13:59            | LB138198      |

## Metals

- 3a -

### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Dal Tile - Sunnyvale Plant

**SDG No.:** Q3796

**Contract:** DALT03

**Lab Code:** ACE

| Sample ID | Analyte    | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|------------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB04     | Aluminum   | 20.3           | +/-50               | J            | 100  | P | 12/11/2025       | 15:04            | LB138198      |
|           | Antimony   | 6.76           | +/-25               | U            | 50.0 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Arsenic    | 5.12           | +/-10               | U            | 20.0 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Barium     | 14.6           | +/-50               | U            | 100  | P | 12/11/2025       | 15:04            | LB138198      |
|           | Beryllium  | 0.56           | +/-3                | U            | 6.00 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Boron      | 15.7           | +/-50               | U            | 100  | P | 12/11/2025       | 15:04            | LB138198      |
|           | Cadmium    | 0.50           | +/-3                | U            | 6.00 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Calcium    | 234            | +/-1000             | U            | 2000 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Chromium   | 2.12           | +/-5                | U            | 10.0 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Cobalt     | 2.26           | +/-15               | U            | 30.0 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Copper     | 4.60           | +/-10               | U            | 20.0 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Iron       | 23.4           | +/-50               | U            | 100  | P | 12/11/2025       | 15:04            | LB138198      |
|           | Lead       | 2.30           | +/-6                | U            | 12.0 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Magnesium  | 244            | +/-1000             | U            | 2000 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Manganese  | 5.94           | +/-10               | U            | 20.0 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Molybdenum | 18.7           | +/-100              | U            | 200  | P | 12/11/2025       | 15:04            | LB138198      |
|           | Nickel     | 3.06           | +/-20               | U            | 40.0 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Potassium  | 918            | +/-1000             | U            | 2000 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Selenium   | 9.64           | +/-10               | U            | 20.0 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Silver     | 1.62           | +/-5                | U            | 10.0 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Sodium     | 868            | +/-1000             | U            | 2000 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Thallium   | 4.38           | +/-20               | U            | 40.0 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Tin        | 5.14           | +/-20               | U            | 40.0 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Titanium   | 6.42           | +/-20               | U            | 40.0 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Vanadium   | 6.26           | +/-20               | U            | 40.0 | P | 12/11/2025       | 15:04            | LB138198      |
|           | Zinc       | 16.7           | +/-20               | U            | 40.0 | P | 12/11/2025       | 15:04            | LB138198      |
| CCB05     | Aluminum   | 13.8           | +/-50               | J            | 100  | P | 12/11/2025       | 15:54            | LB138198      |
|           | Antimony   | 6.76           | +/-25               | U            | 50.0 | P | 12/11/2025       | 15:54            | LB138198      |
|           | Arsenic    | 5.12           | +/-10               | U            | 20.0 | P | 12/11/2025       | 15:54            | LB138198      |
|           | Barium     | 14.6           | +/-50               | U            | 100  | P | 12/11/2025       | 15:54            | LB138198      |
|           | Beryllium  | 0.56           | +/-3                | U            | 6.00 | P | 12/11/2025       | 15:54            | LB138198      |
|           | Boron      | 15.7           | +/-50               | U            | 100  | P | 12/11/2025       | 15:54            | LB138198      |
|           | Cadmium    | 0.50           | +/-3                | U            | 6.00 | P | 12/11/2025       | 15:54            | LB138198      |
|           | Calcium    | 234            | +/-1000             | U            | 2000 | P | 12/11/2025       | 15:54            | LB138198      |
|           | Chromium   | 2.12           | +/-5                | U            | 10.0 | P | 12/11/2025       | 15:54            | LB138198      |
|           | Cobalt     | 2.26           | +/-15               | U            | 30.0 | P | 12/11/2025       | 15:54            | LB138198      |
|           | Copper     | 4.60           | +/-10               | U            | 20.0 | P | 12/11/2025       | 15:54            | LB138198      |
|           | Iron       | 23.4           | +/-50               | U            | 100  | P | 12/11/2025       | 15:54            | LB138198      |
|           | Lead       | 2.30           | +/-6                | U            | 12.0 | P | 12/11/2025       | 15:54            | LB138198      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Dal Tile - Sunnyvale Plant

**SDG No.:** Q3796

**Contract:** DALT03

**Lab Code:** ACE

| Sample ID | Analyte    | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|------------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB05     | Magnesium  | 244            | +/-1000             | U            | 2000 | P | 12/11/2025       | 15:54            | LB138198      |
|           | Manganese  | 5.94           | +/-10               | U            | 20.0 | P | 12/11/2025       | 15:54            | LB138198      |
|           | Molybdenum | 18.7           | +/-100              | U            | 200  | P | 12/11/2025       | 15:54            | LB138198      |
|           | Nickel     | 3.06           | +/-20               | U            | 40.0 | P | 12/11/2025       | 15:54            | LB138198      |
|           | Potassium  | 918            | +/-1000             | U            | 2000 | P | 12/11/2025       | 15:54            | LB138198      |
|           | Selenium   | 9.64           | +/-10               | U            | 20.0 | P | 12/11/2025       | 15:54            | LB138198      |
|           | Silver     | 1.62           | +/-5                | U            | 10.0 | P | 12/11/2025       | 15:54            | LB138198      |
|           | Sodium     | 868            | +/-1000             | U            | 2000 | P | 12/11/2025       | 15:54            | LB138198      |
|           | Thallium   | 4.38           | +/-20               | U            | 40.0 | P | 12/11/2025       | 15:54            | LB138198      |
|           | Tin        | 5.14           | +/-20               | U            | 40.0 | P | 12/11/2025       | 15:54            | LB138198      |
|           | Titanium   | 6.42           | +/-20               | U            | 40.0 | P | 12/11/2025       | 15:54            | LB138198      |
|           | Vanadium   | 6.26           | +/-20               | U            | 40.0 | P | 12/11/2025       | 15:54            | LB138198      |
|           | Zinc       | 16.7           | +/-20               | U            | 40.0 | P | 12/11/2025       | 15:54            | LB138198      |
| CCB06     | Aluminum   | 11.3           | +/-50               | U            | 100  | P | 12/11/2025       | 16:44            | LB138198      |
|           | Antimony   | 6.76           | +/-25               | U            | 50.0 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Arsenic    | 5.12           | +/-10               | U            | 20.0 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Barium     | 14.6           | +/-50               | U            | 100  | P | 12/11/2025       | 16:44            | LB138198      |
|           | Beryllium  | 0.56           | +/-3                | U            | 6.00 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Boron      | 15.7           | +/-50               | U            | 100  | P | 12/11/2025       | 16:44            | LB138198      |
|           | Cadmium    | 0.50           | +/-3                | U            | 6.00 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Calcium    | 234            | +/-1000             | U            | 2000 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Chromium   | 2.12           | +/-5                | U            | 10.0 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Cobalt     | 2.26           | +/-15               | U            | 30.0 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Copper     | 4.60           | +/-10               | U            | 20.0 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Iron       | 23.4           | +/-50               | U            | 100  | P | 12/11/2025       | 16:44            | LB138198      |
|           | Lead       | 2.30           | +/-6                | U            | 12.0 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Magnesium  | 244            | +/-1000             | U            | 2000 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Manganese  | 5.94           | +/-10               | U            | 20.0 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Molybdenum | 18.7           | +/-100              | U            | 200  | P | 12/11/2025       | 16:44            | LB138198      |
|           | Nickel     | 3.06           | +/-20               | U            | 40.0 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Potassium  | 918            | +/-1000             | U            | 2000 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Selenium   | 9.64           | +/-10               | U            | 20.0 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Silver     | 1.62           | +/-5                | U            | 10.0 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Sodium     | 868            | +/-1000             | U            | 2000 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Thallium   | 4.38           | +/-20               | U            | 40.0 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Tin        | 5.14           | +/-20               | U            | 40.0 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Titanium   | 6.42           | +/-20               | U            | 40.0 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Vanadium   | 6.26           | +/-20               | U            | 40.0 | P | 12/11/2025       | 16:44            | LB138198      |
|           | Zinc       | 16.7           | +/-20               | U            | 40.0 | P | 12/11/2025       | 16:44            | LB138198      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal Tile - Sunnyvale Plant

SDG No.: Q3796

Contract: DALT03

Lab Code: ACE

| Sample ID | Analyte    | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|------------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB07     | Aluminum   | 11.3           | +/-50               | U            | 100  | P | 12/11/2025       | 17:57            | LB138198      |
|           | Antimony   | 6.76           | +/-25               | U            | 50.0 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Arsenic    | 5.12           | +/-10               | U            | 20.0 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Barium     | 14.6           | +/-50               | U            | 100  | P | 12/11/2025       | 17:57            | LB138198      |
|           | Beryllium  | 0.56           | +/-3                | U            | 6.00 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Boron      | 15.7           | +/-50               | U            | 100  | P | 12/11/2025       | 17:57            | LB138198      |
|           | Cadmium    | 0.50           | +/-3                | U            | 6.00 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Calcium    | 234            | +/-1000             | U            | 2000 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Chromium   | 2.12           | +/-5                | U            | 10.0 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Cobalt     | 2.26           | +/-15               | U            | 30.0 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Copper     | 4.60           | +/-10               | U            | 20.0 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Iron       | 23.9           | +/-50               | J            | 100  | P | 12/11/2025       | 17:57            | LB138198      |
|           | Lead       | 2.30           | +/-6                | U            | 12.0 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Magnesium  | 244            | +/-1000             | U            | 2000 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Manganese  | 5.94           | +/-10               | U            | 20.0 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Molybdenum | 18.7           | +/-100              | U            | 200  | P | 12/11/2025       | 17:57            | LB138198      |
|           | Nickel     | 3.06           | +/-20               | U            | 40.0 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Potassium  | 918            | +/-1000             | U            | 2000 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Selenium   | 9.64           | +/-10               | U            | 20.0 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Silver     | 1.62           | +/-5                | U            | 10.0 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Sodium     | 868            | +/-1000             | U            | 2000 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Thallium   | 4.38           | +/-20               | U            | 40.0 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Tin        | 5.14           | +/-20               | U            | 40.0 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Titanium   | 6.42           | +/-20               | U            | 40.0 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Vanadium   | 6.26           | +/-20               | U            | 40.0 | P | 12/11/2025       | 17:57            | LB138198      |
|           | Zinc       | 16.7           | +/-20               | U            | 40.0 | P | 12/11/2025       | 17:57            | LB138198      |
| CCB08     | Aluminum   | 11.3           | +/-50               | U            | 100  | P | 12/11/2025       | 18:52            | LB138198      |
|           | Antimony   | 6.76           | +/-25               | U            | 50.0 | P | 12/11/2025       | 18:52            | LB138198      |
|           | Arsenic    | 5.12           | +/-10               | U            | 20.0 | P | 12/11/2025       | 18:52            | LB138198      |
|           | Barium     | 14.6           | +/-50               | U            | 100  | P | 12/11/2025       | 18:52            | LB138198      |
|           | Beryllium  | 0.56           | +/-3                | U            | 6.00 | P | 12/11/2025       | 18:52            | LB138198      |
|           | Boron      | 15.7           | +/-50               | U            | 100  | P | 12/11/2025       | 18:52            | LB138198      |
|           | Cadmium    | 0.50           | +/-3                | U            | 6.00 | P | 12/11/2025       | 18:52            | LB138198      |
|           | Calcium    | 234            | +/-1000             | U            | 2000 | P | 12/11/2025       | 18:52            | LB138198      |
|           | Chromium   | 2.12           | +/-5                | U            | 10.0 | P | 12/11/2025       | 18:52            | LB138198      |
|           | Cobalt     | 2.26           | +/-15               | U            | 30.0 | P | 12/11/2025       | 18:52            | LB138198      |
|           | Copper     | 4.60           | +/-10               | U            | 20.0 | P | 12/11/2025       | 18:52            | LB138198      |
|           | Iron       | 23.4           | +/-50               | U            | 100  | P | 12/11/2025       | 18:52            | LB138198      |
|           | Lead       | 2.30           | +/-6                | U            | 12.0 | P | 12/11/2025       | 18:52            | LB138198      |

## Metals

- 3a -

### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Dal Tile - Sunnyvale Plant

**SDG No.:** Q3796

**Contract:** DALT03

**Lab Code:** ACE

| Sample ID | Analyte    | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|------------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB08     | Magnesium  | 244            | +/-1000             | U            | 2000 | P | 12/11/2025       | 18:52            | LB138198      |
|           | Manganese  | 5.94           | +/-10               | U            | 20.0 | P | 12/11/2025       | 18:52            | LB138198      |
|           | Molybdenum | 18.7           | +/-100              | U            | 200  | P | 12/11/2025       | 18:52            | LB138198      |
|           | Nickel     | 3.06           | +/-20               | U            | 40.0 | P | 12/11/2025       | 18:52            | LB138198      |
|           | Potassium  | 918            | +/-1000             | U            | 2000 | P | 12/11/2025       | 18:52            | LB138198      |
|           | Selenium   | 9.64           | +/-10               | U            | 20.0 | P | 12/11/2025       | 18:52            | LB138198      |
|           | Silver     | 1.62           | +/-5                | U            | 10.0 | P | 12/11/2025       | 18:52            | LB138198      |
|           | Sodium     | 868            | +/-1000             | U            | 2000 | P | 12/11/2025       | 18:52            | LB138198      |
|           | Thallium   | 4.38           | +/-20               | U            | 40.0 | P | 12/11/2025       | 18:52            | LB138198      |
|           | Tin        | 5.14           | +/-20               | U            | 40.0 | P | 12/11/2025       | 18:52            | LB138198      |
|           | Titanium   | 6.42           | +/-20               | U            | 40.0 | P | 12/11/2025       | 18:52            | LB138198      |
|           | Vanadium   | 6.26           | +/-20               | U            | 40.0 | P | 12/11/2025       | 18:52            | LB138198      |
|           | Zinc       | 16.7           | +/-20               | U            | 40.0 | P | 12/11/2025       | 18:52            | LB138198      |
| CCB09     | Aluminum   | 11.3           | +/-50               | U            | 100  | P | 12/11/2025       | 19:45            | LB138198      |
|           | Antimony   | 6.76           | +/-25               | U            | 50.0 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Arsenic    | 5.12           | +/-10               | U            | 20.0 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Barium     | 14.6           | +/-50               | U            | 100  | P | 12/11/2025       | 19:45            | LB138198      |
|           | Beryllium  | 0.56           | +/-3                | U            | 6.00 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Boron      | 15.7           | +/-50               | U            | 100  | P | 12/11/2025       | 19:45            | LB138198      |
|           | Cadmium    | 0.50           | +/-3                | U            | 6.00 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Calcium    | 234            | +/-1000             | U            | 2000 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Chromium   | 2.12           | +/-5                | U            | 10.0 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Cobalt     | 2.26           | +/-15               | U            | 30.0 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Copper     | 4.60           | +/-10               | U            | 20.0 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Iron       | 23.4           | +/-50               | U            | 100  | P | 12/11/2025       | 19:45            | LB138198      |
|           | Lead       | 3.43           | +/-6                | J            | 12.0 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Magnesium  | 244            | +/-1000             | U            | 2000 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Manganese  | 5.94           | +/-10               | U            | 20.0 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Molybdenum | 18.7           | +/-100              | U            | 200  | P | 12/11/2025       | 19:45            | LB138198      |
|           | Nickel     | 3.06           | +/-20               | U            | 40.0 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Potassium  | 918            | +/-1000             | U            | 2000 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Selenium   | 9.64           | +/-10               | U            | 20.0 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Silver     | 1.62           | +/-5                | U            | 10.0 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Sodium     | 868            | +/-1000             | U            | 2000 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Thallium   | 4.38           | +/-20               | U            | 40.0 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Tin        | 5.14           | +/-20               | U            | 40.0 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Titanium   | 6.42           | +/-20               | U            | 40.0 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Vanadium   | 6.26           | +/-20               | U            | 40.0 | P | 12/11/2025       | 19:45            | LB138198      |
|           | Zinc       | 16.7           | +/-20               | U            | 40.0 | P | 12/11/2025       | 19:45            | LB138198      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Dal Tile - Sunnyvale Plant

**SDG No.:** Q3796

**Contract:** DALT03

**Lab Code:** ACE

| Sample ID | Analyte    | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|------------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB10     | Aluminum   | 11.3           | +/-50               | U            | 100  | P | 12/11/2025       | 20:37            | LB138198      |
|           | Antimony   | 6.76           | +/-25               | U            | 50.0 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Arsenic    | 5.12           | +/-10               | U            | 20.0 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Barium     | 14.6           | +/-50               | U            | 100  | P | 12/11/2025       | 20:37            | LB138198      |
|           | Beryllium  | 0.56           | +/-3                | U            | 6.00 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Boron      | 15.7           | +/-50               | U            | 100  | P | 12/11/2025       | 20:37            | LB138198      |
|           | Cadmium    | 0.50           | +/-3                | U            | 6.00 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Calcium    | 234            | +/-1000             | U            | 2000 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Chromium   | 2.12           | +/-5                | U            | 10.0 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Cobalt     | 2.26           | +/-15               | U            | 30.0 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Copper     | 4.60           | +/-10               | U            | 20.0 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Iron       | 23.4           | +/-50               | U            | 100  | P | 12/11/2025       | 20:37            | LB138198      |
|           | Lead       | 2.34           | +/-6                | J            | 12.0 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Magnesium  | 244            | +/-1000             | U            | 2000 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Manganese  | 5.94           | +/-10               | U            | 20.0 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Molybdenum | 18.7           | +/-100              | U            | 200  | P | 12/11/2025       | 20:37            | LB138198      |
|           | Nickel     | 3.06           | +/-20               | U            | 40.0 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Potassium  | 918            | +/-1000             | U            | 2000 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Selenium   | 9.64           | +/-10               | U            | 20.0 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Silver     | 1.62           | +/-5                | U            | 10.0 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Sodium     | 868            | +/-1000             | U            | 2000 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Thallium   | 4.38           | +/-20               | U            | 40.0 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Tin        | 5.14           | +/-20               | U            | 40.0 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Titanium   | 6.42           | +/-20               | U            | 40.0 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Vanadium   | 6.26           | +/-20               | U            | 40.0 | P | 12/11/2025       | 20:37            | LB138198      |
|           | Zinc       | 16.7           | +/-20               | U            | 40.0 | P | 12/11/2025       | 20:37            | LB138198      |
| CCB11     | Aluminum   | 17.9           | +/-50               | J            | 100  | P | 12/11/2025       | 20:59            | LB138198      |
|           | Antimony   | 6.76           | +/-25               | U            | 50.0 | P | 12/11/2025       | 20:59            | LB138198      |
|           | Arsenic    | 5.12           | +/-10               | U            | 20.0 | P | 12/11/2025       | 20:59            | LB138198      |
|           | Barium     | 14.6           | +/-50               | U            | 100  | P | 12/11/2025       | 20:59            | LB138198      |
|           | Beryllium  | 0.56           | +/-3                | U            | 6.00 | P | 12/11/2025       | 20:59            | LB138198      |
|           | Boron      | 15.7           | +/-50               | U            | 100  | P | 12/11/2025       | 20:59            | LB138198      |
|           | Cadmium    | 0.50           | +/-3                | U            | 6.00 | P | 12/11/2025       | 20:59            | LB138198      |
|           | Calcium    | 234            | +/-1000             | U            | 2000 | P | 12/11/2025       | 20:59            | LB138198      |
|           | Chromium   | 2.12           | +/-5                | U            | 10.0 | P | 12/11/2025       | 20:59            | LB138198      |
|           | Cobalt     | 2.26           | +/-15               | U            | 30.0 | P | 12/11/2025       | 20:59            | LB138198      |
|           | Copper     | 4.60           | +/-10               | U            | 20.0 | P | 12/11/2025       | 20:59            | LB138198      |
|           | Iron       | 23.4           | +/-50               | U            | 100  | P | 12/11/2025       | 20:59            | LB138198      |
|           | Lead       | 2.32           | +/-6                | J            | 12.0 | P | 12/11/2025       | 20:59            | LB138198      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Dal Tile - Sunnyvale Plant

**SDG No.:** Q3796

**Contract:** DALT03

**Lab Code:** ACE

| Sample ID | Analyte    | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|------------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB11     | Magnesium  | 244            | +/-1000             | U            | 2000 | P | 12/11/2025       | 20:59            | LB138198      |
|           | Manganese  | 5.94           | +/-10               | U            | 20.0 | P | 12/11/2025       | 20:59            | LB138198      |
|           | Molybdenum | 18.7           | +/-100              | U            | 200  | P | 12/11/2025       | 20:59            | LB138198      |
|           | Nickel     | 3.06           | +/-20               | U            | 40.0 | P | 12/11/2025       | 20:59            | LB138198      |
|           | Potassium  | 918            | +/-1000             | U            | 2000 | P | 12/11/2025       | 20:59            | LB138198      |
|           | Selenium   | 9.64           | +/-10               | U            | 20.0 | P | 12/11/2025       | 20:59            | LB138198      |
|           | Silver     | 1.62           | +/-5                | U            | 10.0 | P | 12/11/2025       | 20:59            | LB138198      |
|           | Sodium     | 868            | +/-1000             | U            | 2000 | P | 12/11/2025       | 20:59            | LB138198      |
|           | Thallium   | 4.38           | +/-20               | U            | 40.0 | P | 12/11/2025       | 20:59            | LB138198      |
|           | Tin        | 5.14           | +/-20               | U            | 40.0 | P | 12/11/2025       | 20:59            | LB138198      |
|           | Titanium   | 6.42           | +/-20               | U            | 40.0 | P | 12/11/2025       | 20:59            | LB138198      |
|           | Vanadium   | 6.26           | +/-20               | U            | 40.0 | P | 12/11/2025       | 20:59            | LB138198      |
|           | Zinc       | 16.7           | +/-20               | U            | 40.0 | P | 12/11/2025       | 20:59            | LB138198      |