

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Dal Tile - Sunnyvale Plant</u>	SDG No.: <u>Q3796</u>
Contract: <u>DALT03</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	249000	255000	98	216000	294000	12/11/2025	11:54	LB138198
	Antimony	4.13			-50	50	12/11/2025	11:54	LB138198
	Arsenic	2.67			-20	20	12/11/2025	11:54	LB138198
	Barium	4.34	6.0	72	-94	106	12/11/2025	11:54	LB138198
	Beryllium	1.04			-6	6	12/11/2025	11:54	LB138198
	Boron	1.62	1000		-100	100	12/11/2025	11:54	LB138198
	Cadmium	3.39	1.0	339	-5	7	12/11/2025	11:54	LB138198
	Calcium	235000	245000	96	208000	282000	12/11/2025	11:54	LB138198
	Chromium	52.4	52.0	101	42	62	12/11/2025	11:54	LB138198
	Cobalt	2.61			-30	30	12/11/2025	11:54	LB138198
	Copper	2.84	2.0	142	-18	22	12/11/2025	11:54	LB138198
	Iron	91200	101000	90	85600	116500	12/11/2025	11:54	LB138198
	Lead	7.82			-12	12	12/11/2025	11:54	LB138198
	Magnesium	254000	255000	100	216000	294000	12/11/2025	11:54	LB138198
	Manganese	9.15	7.0	131	-13	27	12/11/2025	11:54	LB138198
	Molybdenum	1.68	1000		-200	200	12/11/2025	11:54	LB138198
	Nickel	6.93	2.0	346	-38	42	12/11/2025	11:54	LB138198
	Potassium	120			0	0	12/11/2025	11:54	LB138198
	Selenium	1.32			-20	20	12/11/2025	11:54	LB138198
	Silver	7.73			-10	10	12/11/2025	11:54	LB138198
	Sodium	84.2			0	0	12/11/2025	11:54	LB138198
	Thallium	14.0			-40	40	12/11/2025	11:54	LB138198
	Tin	-8.14	1000	1	-40	40	12/11/2025	11:54	LB138198
	Titanium	-4.13			0	0	12/11/2025	11:54	LB138198
	Vanadium	5.97			-40	40	12/11/2025	11:54	LB138198
	Zinc	5.57			-40	40	12/11/2025	11:54	LB138198
ICSAB01	Aluminum	240000	247000	97	209000	285000	12/11/2025	11:59	LB138198
	Antimony	583	618	94	525	711	12/11/2025	11:59	LB138198
	Arsenic	105	104	101	88.4	120	12/11/2025	11:59	LB138198
	Barium	502	537	94	437	637	12/11/2025	11:59	LB138198
	Beryllium	485	495	98	420	570	12/11/2025	11:59	LB138198
	Boron	930	1000	93	850	1150	12/11/2025	11:59	LB138198
	Cadmium	946	972	97	826	1120	12/11/2025	11:59	LB138198
	Calcium	229000	235000	97	199000	271000	12/11/2025	11:59	LB138198
	Chromium	535	542	99	460	624	12/11/2025	11:59	LB138198
	Cobalt	475	476	100	404	548	12/11/2025	11:59	LB138198
	Copper	458	511	90	434	588	12/11/2025	11:59	LB138198
	Iron	89800	99300	90	84400	114500	12/11/2025	11:59	LB138198
	Lead	58.0	49.0	118	37	61	12/11/2025	11:59	LB138198
	Magnesium	243000	248000	98	210000	286000	12/11/2025	11:59	LB138198
	Manganese	494	507	97	430	584	12/11/2025	11:59	LB138198
	Molybdenum	924	1000	92	850	1150	12/11/2025	11:59	LB138198

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Dal Tile - Sunnyvale Plant

SDG No.: Q3796

Contract: DALT03

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Nickel	944	954	99	810	1100	12/11/2025	11:59	LB138198
	Potassium	58.7			0	0	12/11/2025	11:59	LB138198
	Selenium	46.2	46.0	100	26	66	12/11/2025	11:59	LB138198
	Silver	180	201	90	170	232	12/11/2025	11:59	LB138198
	Sodium	100			0	0	12/11/2025	11:59	LB138198
	Thallium	119	108	110	68	148	12/11/2025	11:59	LB138198
	Tin	943	1000	94	850	1150	12/11/2025	11:59	LB138198
	Titanium	935			0	0	12/11/2025	11:59	LB138198
	Vanadium	470	491	96	417	565	12/11/2025	11:59	LB138198
	Zinc	987	952	104	809	1095	12/11/2025	11:59	LB138198