

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Dal Tile - Sunnyvale Plant

SDG No.: Q3796

Contract: DALT03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170849BS							
Aluminum	ug/L	1000	989		99	80 - 120	P
Antimony	ug/L	400	394		98	80 - 120	P
Arsenic	ug/L	400	390		98	80 - 120	P
Barium	ug/L	100	102		102	80 - 120	P
Beryllium	ug/L	100	102		102	80 - 120	P
Boron	ug/L	150	152		101	80 - 120	P
Cadmium	ug/L	100	95.2		95	80 - 120	P
Calcium	ug/L	500	508	J	102	80 - 120	P
Chromium	ug/L	200	206		103	80 - 120	P
Cobalt	ug/L	100	95.9		96	80 - 120	P
Copper	ug/L	150	151		101	80 - 120	P
Iron	ug/L	1500	1460		97	80 - 120	P
Lead	ug/L	500	472		94	80 - 120	P
Magnesium	ug/L	1000	1000		100	80 - 120	P
Manganese	ug/L	100	103		103	80 - 120	P
Molybdenum	ug/L	200	199		100	80 - 120	P
Nickel	ug/L	250	242		97	80 - 120	P
Potassium	ug/L	5000	4790		96	80 - 120	P
Selenium	ug/L	1000	991		99	80 - 120	P
Silver	ug/L	37.5	36.7		98	80 - 120	P
Sodium	ug/L	1500	1500		100	80 - 120	P
Thallium	ug/L	1000	1040		104	80 - 120	P
Tin	ug/L	350	333		95	80 - 120	P
Titanium	ug/L	100	100		100	80 - 120	P
Vanadium	ug/L	150	150		100	80 - 120	P
Zinc	ug/L	100	106		106	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Dal Tile - Sunnyvale Plant

SDG No.: Q3796

Contract: DALT03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170852BS Mercury	ug/L	4.0	3.51		88	82 - 119	CV